

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110618\
 Data File : VX005817.D
 Acq On : 06 Nov 2018 18:51
 Operator : JC/MD
 Sample : J5765-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NB-08-110518

Quant Time: Nov 07 09:34:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	162204	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	274407	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	236373	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	103246	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	116925	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
35) Dibromofluoromethane	5.50	113	92007	48.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.84%	
50) Toluene-d8	8.71	98	322693	48.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.84%	
62) 4-Bromofluorobenzene	11.14	95	102852	41.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.64%	
Target Compounds						
18) Methyl Acetate	2.78	43	7009	2.059	ug/l #	90
20) Methylene Chloride	2.86	84	4166	2.037	ug/l	88
43) Isopropyl Acetate	6.46	43	121981	22.532	ug/l	97
95) Naphthalene	13.83	128	111498	15.487	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110618\
Data File : VX005817.D
Acq On : 06 Nov 2018 18:51
Operator : JC/MD
Sample : J5765-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NB-08-110518

Quant Time: Nov 07 09:34:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

