

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110619\
 Data File : VX013337.D
 Acq On : 06 Nov 2019 12:18
 Operator : JC/SP
 Sample : K5711-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-191105

Quant Time: Nov 07 02:33:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	778271	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1190017	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1063181	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	489769	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	436833	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
35) Dibromofluoromethane	5.48	113	360245	48.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.52%	
50) Toluene-d8	8.71	98	1417016	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
62) 4-Bromofluorobenzene	11.13	95	479401	47.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.42%	
Target Compounds						
3) Chloromethane	1.32	50	7000	0.720	ug/l	91
16) Acetone	2.43	43	13093	2.847	ug/l	100
27) cis-1,2-Dichloroethene	4.58	96	33332	3.828	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110619\
Data File : VX013337.D
Acq On : 06 Nov 2019 12:18
Operator : JC/SP
Sample : K5711-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-191105

Quant Time: Nov 07 02:33:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

