

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110619\
 Data File : VX013339.D
 Acq On : 06 Nov 2019 13:04
 Operator : JC/SP
 Sample : K5713-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-TB-20191105

Quant Time: Nov 07 02:36:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	780484	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1190991	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1047433	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	483298	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	430177	46.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.04%	
35) Dibromofluoromethane	5.48	113	359262	48.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.18%	
50) Toluene-d8	8.71	98	1420713	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
62) 4-Bromofluorobenzene	11.13	95	470892	46.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.66%	
Target Compounds						
16) Acetone	2.43	43	55323	11.995	ug/l	97
20) Methylene Chloride	2.84	84	9520	1.117	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110619\
Data File : VX013339.D
Acq On : 06 Nov 2019 13:04
Operator : JC/SP
Sample : K5713-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20191105

Quant Time: Nov 07 02:36:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

