

Data File : VX019283.D
Acq On : 04 Nov 2020 20:49
Operator : JC/MD
Sample : L4616-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA20-CHY07

Quant Time: Nov 05 07:14:51 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	281850	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	501868	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	486602	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	242411	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	200562	55.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.36%	
35) Dibromofluoromethane	5.46	113	157033	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
50) Toluene-d8	8.70	98	626068	52.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.36%	
62) 4-Bromofluorobenzene	11.12	95	239970	53.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.64%	
Target Compounds						
16) Acetone	2.42	43	4019	3.205	ug/l	97
17) Carbon Disulfide	2.55	76	1234	0.916	ug/l #	88
20) Methylene Chloride	2.84	84	1051	0.334	ug/l #	82
44) Trichloroethene	7.19	130	1956	0.520	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX019283.D
Acq On : 04 Nov 2020 20:49
Operator : JC/MD
Sample : L4616-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
FA20-CHY07

Quant Time: Nov 05 07:14:51 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

