

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110620\
 Data File : VX019285.D
 Acq On : 04 Nov 2020 21:33
 Operator : JC/MD
 Sample : L4616-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA20-MMW3414

Quant Time: Nov 21 06:38:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	288338	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	507791	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	484816	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	240076	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	201992	54.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.64%	
35) Dibromofluoromethane	5.46	113	157743	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
50) Toluene-d8	8.70	98	633375	52.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.36%	
62) 4-Bromofluorobenzene	11.12	95	238548	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
Target Compounds						
16) Acetone	2.42	43	3745	2.919	ug/l	97
17) Carbon Disulfide	2.56	76	1277	0.918	ug/l #	84
27) cis-1,2-Dichloroethene	4.57	96	1395	0.388	ug/l	90
44) Trichloroethene	7.19	130	1311	0.344	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110620\
Data File : VX019285.D
Acq On : 04 Nov 2020 21:33
Operator : JC/MD
Sample : L4616-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA20-MMW3414

Quant Time: Nov 21 06:38:13 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

