

Data File : VX019286.D
Acq On : 04 Nov 2020 21:56
Operator : JC/MD
Sample : L4616-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA20-CHY02

Quant Time: Nov 05 03:11:14 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	276023	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	488055	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	466324	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	233115	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	196307	55.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.30%	
35) Dibromofluoromethane	5.46	113	154220	50.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.62%	
50) Toluene-d8	8.70	98	608305	52.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.28%	
62) 4-Bromofluorobenzene	11.12	95	233823	53.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.86%	

Target Compounds

						Qvalue
3) Chloromethane	1.31	50	1341	0.522	ug/l	98
16) Acetone	2.42	43	11416	9.296	ug/l	94
17) Carbon Disulfide	2.55	76	2317	1.059	ug/l	98
20) Methylene Chloride	2.83	84	1147	0.372	ug/l #	91
25) 2-Butanone	4.65	43	3954	2.043	ug/l	96
27) cis-1,2-Dichloroethene	4.57	96	2295	0.667	ug/l	97
44) Trichloroethene	7.19	130	2440	0.667	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX019286.D
Acq On : 04 Nov 2020 21:56
Operator : JC/MD
Sample : L4616-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleID :
FA20-CHY02

Quant Time: Nov 05 03:11:14 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

