

Data File : VX019287.D
Acq On : 04 Nov 2020 22:18
Operator : JC/MD
Sample : L4616-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA20-CHY08

Quant Time: Nov 05 07:20:34 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	288054	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	512597	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	494008	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	251949	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	206022	55.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.92%	
35) Dibromofluoromethane	5.46	113	161352	50.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.22%	
50) Toluene-d8	8.70	98	641641	52.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.72%	
62) 4-Bromofluorobenzene	11.12	95	240948	52.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.84%	

Target Compounds

						Qvalue
3) Chloromethane	1.31	50	1306	0.487	ug/l	94
16) Acetone	2.42	43	4251	3.317	ug/l	97
17) Carbon Disulfide	2.56	76	509	0.823	ug/l #	57
20) Methylene Chloride	2.84	84	1538	0.478	ug/l	90
44) Trichloroethene	7.18	130	1378	0.359	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX019287.D
Acq On : 04 Nov 2020 22:18
Operator : JC/MD
Sample : L4616-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA20-CHY08

Quant Time: Nov 05 07:20:34 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

