

Data File : VX019289.D  
Acq On : 04 Nov 2020 23:03  
Operator : JC/MD  
Sample : L4607-02  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
OW-01

Quant Time: Nov 05 07:24:14 2020  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X102720W.M  
Quant Title : SW846 8260  
QLast Update : Tue Oct 27 13:55:07 2020  
Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|--------|---------|----------|
| 1) Pentafluorobenzene       | 5.64  | 168  | 281866   | 50.00  | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.83  | 114  | 493244   | 50.00  | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 10.10 | 117  | 470933   | 50.00  | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.06 | 152  | 234731   | 50.00  | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |        |         |          |
| 33) 1,2-Dichloroethane-d4   | 6.03  | 65   | 198388   | 54.58  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 109.16% |          |
| 35) Dibromofluoromethane    | 5.46  | 113  | 158480   | 51.15  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 102.30% |          |
| 50) Toluene-d8              | 8.70  | 98   | 616439   | 52.78  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 105.56% |          |
| 62) 4-Bromofluorobenzene    | 11.12 | 95   | 233111   | 52.70  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 105.40% |          |
| Target Compounds            |       |      |          |        |         |          |
| 11) Tert butyl alcohol      | 3.01  | 59   | 16064    | 22.374 | ug/l #  | 94       |
| 16) Acetone                 | 2.42  | 43   | 7775     | 6.200  | ug/l    | 99       |
| 17) Carbon Disulfide        | 2.55  | 76   | 2074     | 1.022  | ug/l    | 100      |
| 24) 1,1-Dichloroethane      | 3.67  | 63   | 4362     | 0.847  | ug/l    | 96       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX019289.D  
Acq On : 04 Nov 2020 23:03  
Operator : JC/MD  
Sample : L4607-02  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
OW-01

Quant Time: Nov 05 07:24:14 2020  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X102720W.M  
Quant Title : SW846 8260  
QLast Update : Tue Oct 27 13:55:07 2020  
Response via : Initial Calibration

