

Data File : VX019290.D
Acq On : 04 Nov 2020 23:25
Operator : JC/MD
Sample : L4607-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-02

Quant Time: Nov 05 07:25:32 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	279644	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	501716	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	482558	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	237238	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	200015	55.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.92%	
35) Dibromofluoromethane	5.46	113	156835	49.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.54%	
50) Toluene-d8	8.70	98	627898	52.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.70%	
62) 4-Bromofluorobenzene	11.12	95	238566	53.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.06%	
Target Compounds						
16) Acetone	2.42	43	4739	3.809	ug/l	97
31) Cyclohexane	5.56	56	2031	0.446	ug/l	90
39) Methylcyclohexane	7.44	83	2505	0.452	ug/l #	92
95) Naphthalene	13.82	128	8718	0.494	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX019290.D
Acq On : 04 Nov 2020 23:25
Operator : JC/MD
Sample : L4607-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-02

Quant Time: Nov 05 07:25:32 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

