

Data File : VX019292.D
Acq On : 05 Nov 2020 00:10
Operator : JC/MD
Sample : L4607-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-01

Quant Time: Nov 05 03:26:49 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	282744	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	501250	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	471365	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	239806	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	201297	55.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.42%	
35) Dibromofluoromethane	5.46	113	160247	50.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.80%	
50) Toluene-d8	8.70	98	627080	52.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.66%	
62) 4-Bromofluorobenzene	11.12	95	235674	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
Target Compounds						
16) Acetone	2.42	43	5526	4.393	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX019292.D
Acq On : 05 Nov 2020 00:10
Operator : JC/MD
Sample : L4607-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-01

Quant Time: Nov 05 03:26:49 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

