

Data File : VX019296.D
Acq On : 05 Nov 2020 01:40
Operator : JC/MD
Sample : L4622-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAV-B7-110220-0-1

Quant Time: Nov 05 03:34:50 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	291648	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	512733	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	499364	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	248563	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	208100	55.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.66%	
35) Dibromofluoromethane	5.46	113	164066	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
50) Toluene-d8	8.70	98	647249	53.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.62%	
62) 4-Bromofluorobenzene	11.12	95	244955	53.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.56%	
Target Compounds						
16) Acetone	2.42	43	14462	11.146	ug/l	99
20) Methylene Chloride	2.84	84	6038	1.854	ug/l	95
43) Isopropyl Acetate	6.41	43	73575	9.995	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX019296.D
Acq On : 05 Nov 2020 01:40
Operator : JC/MD
Sample : L4622-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAV-B7-110220-0-1

Quant Time: Nov 05 03:34:50 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

