

Data File : VX019297.D
Acq On : 05 Nov 2020 02:02
Operator : JC/MD
Sample : L4622-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAV-B7-110220-DP

Quant Time: Nov 05 03:36:22 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	289685	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	517518	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	499151	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	243494	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	207789	55.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.24%	
35) Dibromofluoromethane	5.46	113	166305	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
50) Toluene-d8	8.70	98	652410	53.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.48%	
62) 4-Bromofluorobenzene	11.12	95	244283	52.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.28%	
Target Compounds						
16) Acetone	2.42	43	15891	12.330	ug/l	97
20) Methylene Chloride	2.84	84	6369	1.968	ug/l	97
43) Isopropyl Acetate	6.41	43	73003	9.826	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX019297.D
Acq On : 05 Nov 2020 02:02
Operator : JC/MD
Sample : L4622-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAV-B7-110220-DP

Quant Time: Nov 05 03:36:22 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

