

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110620\
 Data File : VX019301.D
 Acq On : 05 Nov 2020 10:15
 Operator : JC/MD
 Sample : VX1106WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1106WBL02

Quant Time: Nov 06 03:54:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	310273	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.82	114	543284	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	499451	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	238093	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	214578	53.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.26%	
35) Dibromofluoromethane	5.46	113	172414	50.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.04%	
50) Toluene-d8	8.70	98	661855	51.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.90%	
62) 4-Bromofluorobenzene	11.12	95	240986	49.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.94%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110620\
Data File : VX019301.D
Acq On : 05 Nov 2020 10:15
Operator : JC/MD
Sample : VX1106WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX1106WBL02

Quant Time: Nov 06 03:54:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

