

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110620\
 Data File : VX019304.D
 Acq On : 05 Nov 2020 11:33
 Operator : JC/MD
 Sample : L4621-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 06 01:24:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	285235	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	505058	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	474152	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	229443	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	204890	55.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.40%	
35) Dibromofluoromethane	5.46	113	159915	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
50) Toluene-d8	8.70	98	622904	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
62) 4-Bromofluorobenzene	11.12	95	231462	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110620\
Data File : VX019304.D
Acq On : 05 Nov 2020 11:33
Operator : JC/MD
Sample : L4621-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 06 01:24:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

