

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110620\
 Data File : VX019305.D
 Acq On : 05 Nov 2020 11:56
 Operator : JC/MD
 Sample : L4621-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Nov 06 07:28:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	279139	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	490508	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	452770	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	222188	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	197499	54.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.72%	
35) Dibromofluoromethane	5.46	113	155255	50.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.78%	
50) Toluene-d8	8.70	98	596564	51.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.72%	
62) 4-Bromofluorobenzene	11.12	95	223815	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
Target Compounds						
3) Chloromethane	1.31	50	1321	0.508	ug/l	93
11) Tert butyl alcohol	3.00	59	3628	5.103	ug/l #	80
16) Acetone	2.42	43	8192	6.596	ug/l	100
20) Methylene Chloride	2.84	84	1277	0.410	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110620\
Data File : VX019305.D
Acq On : 05 Nov 2020 11:56
Operator : JC/MD
Sample : L4621-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Nov 06 07:28:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

