

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110620\
 Data File : VX019306.D
 Acq On : 05 Nov 2020 12:18
 Operator : JC/MD
 Sample : L4612-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 265316

Quant Time: Nov 06 07:31:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	276415	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	493247	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	463586	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	228733	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	200824	56.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.68%	
35) Dibromofluoromethane	5.46	113	155913	50.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.64%	
50) Toluene-d8	8.70	98	601850	51.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.06%	
62) 4-Bromofluorobenzene	11.12	95	230031	52.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.02%	
Target Compounds						
16) Acetone	2.42	43	66350	53.953	ug/l	97
18) Methyl Acetate	2.75	43	26368	8.817	ug/l	97
25) 2-Butanone	4.65	43	9372	4.836	ug/l	92
29) Tetrahydrofuran	5.10	42	8263	6.765	ug/l	92
84) 1,2,4-Trimethylbenzene	11.79	105	4606	0.334	ug/l	98
95) Naphthalene	13.82	128	12684	0.746	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110620\
Data File : VX019306.D
Acq On : 05 Nov 2020 12:18
Operator : JC/MD
Sample : L4612-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
265316

Quant Time: Nov 06 07:31:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

