

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110620\
 Data File : VX019307.D
 Acq On : 05 Nov 2020 12:41
 Operator : JC/MD
 Sample : L4634-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6

Quant Time: Nov 06 07:34:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	282204	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	499798	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	466587	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	238987	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	201087	55.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.50%	
35) Dibromofluoromethane	5.46	113	157008	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
50) Toluene-d8	8.70	98	616169	52.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.14%	
62) 4-Bromofluorobenzene	11.12	95	235861	52.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.26%	
Target Compounds						
3) Chloromethane	1.31	50	2380	0.906	ug/l	96
7) Trichlorofluoromethane	1.92	101	7879	1.645	ug/l	94
12) 1,1-Dichloroethene	2.36	96	1516	0.554	ug/l #	83
16) Acetone	2.42	43	11249	8.960	ug/l	94
20) Methylene Chloride	2.83	84	1364	0.433	ug/l	90
27) cis-1,2-Dichloroethene	4.57	96	4277	1.216	ug/l	87
44) Trichloroethene	7.19	130	22906	6.112	ug/l	93
64) Tetrachloroethene	9.32	164	160829	45.242	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110620\
Data File : VX019307.D
Acq On : 05 Nov 2020 12:41
Operator : JC/MD
Sample : L4634-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6

Quant Time: Nov 06 07:34:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

