

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110620\
 Data File : VX019309.D
 Acq On : 05 Nov 2020 13:26
 Operator : JC/MD
 Sample : PB132825TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB132825TB

Quant Time: Nov 06 01:42:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	259832	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	461101	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	442655	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	211458	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	187721	56.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.04%	
35) Dibromofluoromethane	5.47	113	146893	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
50) Toluene-d8	8.70	98	572391	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
62) 4-Bromofluorobenzene	11.12	95	219150	53.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.00%	
Target Compounds						
16) Acetone	2.42	43	8348	7.222	ug/l #	86
18) Methyl Acetate	2.75	43	2956	1.052	ug/l	93
43) Isopropyl Acetate	6.41	43	77853	11.761	ug/l	98
64) Tetrachloroethene	9.32	164	3549	1.052	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110620\
Data File : VX019309.D
Acq On : 05 Nov 2020 13:26
Operator : JC/MD
Sample : PB132825TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB132825TB

Quant Time: Nov 06 01:42:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

