

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110620\
 Data File : VX019312.D
 Acq On : 05 Nov 2020 14:33
 Operator : JC/MD
 Sample : PB132824ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB132824ZHE#02

Quant Time: Nov 06 01:53:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	251384	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	455525	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	436469	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	214374	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	183737	56.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.36%	
35) Dibromofluoromethane	5.46	113	146102	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
50) Toluene-d8	8.70	98	565959	52.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.94%	
62) 4-Bromofluorobenzene	11.12	95	217138	53.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.32%	
Target Compounds						
16) Acetone	2.42	43	15233	13.620	ug/l	97
18) Methyl Acetate	2.75	43	3071	1.129	ug/l	95
20) Methylene Chloride	2.84	84	7377	2.627	ug/l	96
43) Isopropyl Acetate	6.41	43	76568	11.708	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110620\
Data File : VX019312.D
Acq On : 05 Nov 2020 14:33
Operator : JC/MD
Sample : PB132824ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB132824ZHE#02

Quant Time: Nov 06 01:53:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

