

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110620\
 Data File : VX019313.D
 Acq On : 05 Nov 2020 14:55
 Operator : JC/MD
 Sample : PB132824ZHE#03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB132824ZHE#03

Quant Time: Nov 06 01:55:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	263424	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	473489	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	452304	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	221094	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	191738	56.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.88%	
35) Dibromofluoromethane	5.46	113	151707	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
50) Toluene-d8	8.70	98	586481	52.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.62%	
62) 4-Bromofluorobenzene	11.12	95	224134	52.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.58%	
Target Compounds						
16) Acetone	2.42	43	15632	13.338	ug/l	90
18) Methyl Acetate	2.75	43	3014	1.058	ug/l	97
20) Methylene Chloride	2.84	84	7599	2.583	ug/l	95
43) Isopropyl Acetate	6.41	43	77477	11.398	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110620\
Data File : VX019313.D
Acq On : 05 Nov 2020 14:55
Operator : JC/MD
Sample : PB132824ZHE#03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB132824ZHE#03

Quant Time: Nov 06 01:55:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

