

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110620\
 Data File : VX019317.D
 Acq On : 05 Nov 2020 16:25
 Operator : JC/MD
 Sample : PB132824ZHE#07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB132824ZHE#07

Quant Time: Nov 06 02:05:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	272462	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	484741	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	467950	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	228513	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	197491	56.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.42%	
35) Dibromofluoromethane	5.46	113	156364	51.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.70%	
50) Toluene-d8	8.70	98	606538	52.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.68%	
62) 4-Bromofluorobenzene	11.12	95	230271	52.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.96%	
Target Compounds						
16) Acetone	2.42	43	10113	8.343	ug/l	94
20) Methylene Chloride	2.84	84	6282	2.064	ug/l	98
43) Isopropyl Acetate	6.41	43	77447	11.129	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110620\
Data File : VX019317.D
Acq On : 05 Nov 2020 16:25
Operator : JC/MD
Sample : PB132824ZHE#07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB132824ZHE#07

Quant Time: Nov 06 02:05:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

