

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110620\
 Data File : VX019300.D
 Acq On : 05 Nov 2020 09:41
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 06 01:14:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	0.417	0.427	-2.4	92	0.00
3 P	Chloromethane	0.466	0.470	-0.9	93	0.00
4 C	Vinyl Chloride	0.540	0.571	-5.7#	95	0.00
5 T	Bromomethane	0.369	0.345	6.5	85	0.00
6 T	Chloroethane	0.345	0.376	-9.0	96	0.00
7 T	Trichlorofluoromethane	0.848	0.886	-4.5	91	0.00
8 T	Diethyl Ether	0.326	0.346	-6.1	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.470	0.504	-7.2	95	0.00
10 T	Methyl Iodide	0.478	0.470	1.7	81	0.00
11 T	Tert butyl alcohol	0.127	0.117	7.9	84	0.00
12 CM	1,1-Dichloroethene	0.485	0.488	-0.6#	91	0.00
13 T	Acrolein	0.062	0.066	-6.5	107	0.00
14 T	Allyl chloride	0.734	0.789	-7.5	95	0.00
15 T	Acrylonitrile	0.268	0.279	-4.1	90	0.00
16 T	Acetone	0.222	0.250	-12.6	100	0.00
17 T	Carbon Disulfide	1.473	1.352	8.2	89	0.00
18 T	Methyl Acetate	0.541	0.574	-6.1	95	0.00
19 T	Methyl tert-butyl Ether	1.679	1.766	-5.2	91	0.00
20 T	Methylene Chloride	0.558	0.566	-1.4	92	0.00
21 T	trans-1,2-Dichloroethene	0.558	0.561	-0.5	91	0.00
22 T	Diisopropyl ether	1.463	1.642	-12.2	97	0.00
23 T	Vinyl Acetate	1.305	1.443	-10.6	94	0.00
24 P	1,1-Dichloroethane	0.914	0.994	-8.8	95	0.00
25 T	2-Butanone	0.351	0.382	-8.8	94	0.00
26 T	2,2-Dichloropropane	0.859	0.912	-6.2	93	0.00
27 T	cis-1,2-Dichloroethene	0.623	0.641	-2.9	91	0.00
28 T	Bromochloromethane	0.415	0.430	-3.6	99	0.00
29 T	Tetrahydrofuran	0.221	0.235	-6.3	89	0.00
30 C	Chloroform	0.973	1.037	-6.6#	92	0.00
31 T	Cyclohexane	0.815	0.872	-7.0	95	0.00
32 T	1,1,1-Trichloroethane	0.882	0.932	-5.7	92	0.00
33 S	1,2-Dichloroethane-d4	0.645	0.635	1.6	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.314	0.301	4.1	93	0.00
36 T	1,1-Dichloropropene	0.462	0.483	-4.5	94	0.00
37 T	Ethyl Acetate	0.429	0.442	-3.0	90	0.00
38 T	Carbon Tetrachloride	0.471	0.490	-4.0	91	0.00
39 T	Methylcyclohexane	0.552	0.585	-6.0	93	0.00
40 TM	Benzene	1.309	1.388	-6.0	93	0.00
41 T	Methacrylonitrile	0.231	0.243	-5.2	92	0.00
42 TM	1,2-Dichloroethane	0.482	0.511	-6.0	93	0.00
43 T	Isopropyl Acetate	0.718	0.740	-3.1	90	0.00
44 TM	Trichloroethene	0.375	0.371	1.1	89	0.00
45 C	1,2-Dichloropropane	0.321	0.353	-10.0#	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX110620\
 Data File : VX019300.D
 Acq On : 05 Nov 2020 09:41
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 06 01:14:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.237	0.241	-1.7	91	0.00
47 T	Bromodichloromethane	0.467	0.496	-6.2	91	0.00
48 T	Methyl methacrylate	0.345	0.365	-5.8	91	0.00
49 T	1,4-Dioxane	0.008	0.007	12.5	83	0.00
50 S	Toluene-d8	1.184	1.154	2.5	94	0.00
51 T	4-Methyl-2-Pentanone	0.413	0.443	-7.3	91	0.00
52 CM	Toluene	0.843	0.889	-5.5#	93	0.00
53 T	t-1,3-Dichloropropene	0.534	0.550	-3.0	91	0.00
54 T	cis-1,3-Dichloropropene	0.561	0.594	-5.9	92	0.00
55 T	1,1,2-Trichloroethane	0.333	0.345	-3.6	91	0.00
56 T	Ethyl methacrylate	0.503	0.528	-5.0	88	0.00
57 T	1,3-Dichloropropane	0.562	0.587	-4.4	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.281	0.263	6.4	85	0.00
59 T	2-Hexanone	0.314	0.348	-10.8	94	0.00
60 T	Dibromochloromethane	0.362	0.377	-4.1	88	0.00
61 T	1,2-Dibromoethane	0.356	0.362	-1.7	90	0.00
62 S	4-Bromofluorobenzene	0.448	0.440	1.8	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.381	0.379	0.5	93	0.00
65 PM	Chlorobenzene	1.022	1.012	1.0	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.368	0.370	-0.5	90	0.00
67 C	Ethyl Benzene	1.821	1.845	-1.3#	92	0.00
68 T	m/p-Xylenes	0.683	0.694	-1.6	91	0.00
69 T	o-Xylene	0.648	0.653	-0.8	91	0.00
70 T	Styrene	1.106	1.136	-2.7	92	0.00
71 P	Bromoform	0.292	0.284	2.7	85	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.539	3.477	1.8	92	0.00
74 T	N-amyl acetate	1.294	1.318	-1.9	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.106	1.087	1.7	91	0.00
76 T	1,2,3-Trichloropropane	1.032	1.042	-1.0	94	0.00
77 T	Bromobenzene	0.878	0.836	4.8	89	0.00
78 T	n-propylbenzene	4.040	4.053	-0.3	94	0.00
79 T	2-Chlorotoluene	2.379	2.402	-1.0	95	0.00
80 T	1,3,5-Trimethylbenzene	2.990	2.989	0.0	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.392	0.373	4.8	89	0.00
82 T	4-Chlorotoluene	2.905	2.856	1.7	95	0.00
83 T	tert-Butylbenzene	2.863	2.827	1.3	92	0.00
84 T	1,2,4-Trimethylbenzene	3.019	3.006	0.4	93	0.00
85 T	sec-Butylbenzene	3.434	3.479	-1.3	95	0.00
86 T	p-Isopropyltoluene	3.169	3.216	-1.5	94	0.00
87 T	1,3-Dichlorobenzene	1.636	1.568	4.2	92	0.00
88 T	1,4-Dichlorobenzene	1.703	1.596	6.3	93	0.00
89 T	n-Butylbenzene	2.979	3.049	-2.3	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110620\
Data File : VX019300.D
Acq On : 05 Nov 2020 09:41
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 06 01:14:25 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.539	0.536	0.6	90	0.00
91 T	1,2-Dichlorobenzene	1.571	1.477	6.0	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.271	0.247	8.9	84	0.00
93 T	1,2,4-Trichlorobenzene	1.180	1.074	9.0	90	0.00
94 T	Hexachlorobutadiene	0.519	0.467	10.0	90	0.00
95 T	Naphthalene	3.716	3.393	8.7	88	0.00
96 T	1,2,3-Trichlorobenzene	1.145	1.037	9.4	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6