

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110718\
 Data File : VX005829.D
 Acq On : 07 Nov 2018 15:54
 Operator : JC/MD
 Sample : VX1107WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1107WBL01

Quant Time: Nov 07 16:24:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	117053	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	169600	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	136396	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	68224	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	71490	53.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.26%	
35) Dibromofluoromethane	5.50	113	55862	48.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.86%	
50) Toluene-d8	8.71	98	179259	46.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.48%	
62) 4-Bromofluorobenzene	11.13	95	61194	44.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.76%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110718\
Data File : VX005829.D
Acq On : 07 Nov 2018 15:54
Operator : JC/MD
Sample : VX1107WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1107WBL01

Quant Time: Nov 07 16:24:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

