

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110718\
 Data File : VX005834.D
 Acq On : 07 Nov 2018 18:11
 Operator : JC/MD
 Sample : J5848-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S81-0100

Quant Time: Nov 08 10:44:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	121693	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	154661	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	132987	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	68762	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	69294	49.53	ug/l	0.00
Spiked Amount			Recovery	=	99.06%	
35) Dibromofluoromethane	5.50	113	57665	54.82	ug/l	0.00
Spiked Amount			Recovery	=	109.64%	
50) Toluene-d8	8.71	98	182672	52.23	ug/l	0.00
Spiked Amount			Recovery	=	104.46%	
62) 4-Bromofluorobenzene	11.14	95	63924	50.84	ug/l	0.00
Spiked Amount			Recovery	=	101.68%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	21435	28.456	ug/l	98
17) Carbon Disulfide	2.58	76	4996	1.557	ug/l	100
25) 2-Butanone	4.70	43	2239	1.891	ug/l #	89
27) cis-1,2-Dichloroethene	4.59	96	2672	2.133	ug/l	96
31) Cyclohexane	5.58	56	14170	7.271	ug/l	98
39) Methylcyclohexane	7.46	83	17181	11.370	ug/l	94
40) Benzene	6.14	78	5502	1.266	ug/l	95
67) Ethyl Benzene	10.25	91	40677	9.522	ug/l	99
68) m/p-Xylenes	10.36	106	5203	3.193	ug/l	96
69) o-Xylene	10.70	106	41630	28.490	ug/l	99
73) Isopropylbenzene	11.02	105	4222	1.062	ug/l	100
78) n-propylbenzene	11.36	91	5664	1.209	ug/l	98
84) 1,2,4-Trimethylbenzene	11.81	105	24086	6.354	ug/l	99
95) Naphthalene	13.83	128	7492	2.656	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110718\
Data File : VX005834.D
Acq On : 07 Nov 2018 18:11
Operator : JC/MD
Sample : J5848-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S81-0100

Quant Time: Nov 08 10:44:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

