


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110718\
Data File : VX005821.D
Acq On : 07 Nov 2018 11:31
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Title : SW846 8260
Last Update : Wed Nov 07 14:40:53 2018



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.4	2941	PASS
75	95	30	60	48.2	7691	PASS
95	95	100	100	100.0	15963	PASS
96	95	5	9	6.8	1080	PASS
173	174	0.00	2	1.1	160	PASS
174	95	50	100	88.8	14175	PASS
175	174	5	9	7.1	1001	PASS
176	174	95	101	95.1	13475	PASS
177	176	5	9	6.8	916	PASS