

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110719\
 Data File : VX013349.D
 Acq On : 07 Nov 2019 16:16
 Operator : JC/SP
 Sample : K5626-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS003RW114R-191030

Quant Time: Nov 07 16:52:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	809677	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1258019	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1113546	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	532875	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	457686	47.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.42%	
35) Dibromofluoromethane	5.48	113	377964	47.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.80%	
50) Toluene-d8	8.71	98	1489087	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
62) 4-Bromofluorobenzene	11.13	95	509934	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
Target Compounds						
16) Acetone	2.42	43	14824	3.098	ug/l	92
40) Benzene	6.14	78	9132	0.272	ug/l #	88
65) Chlorobenzene	10.13	112	724798	32.150	ug/l	99
87) 1,3-Dichlorobenzene	12.02	146	177624	10.580	ug/l	98
88) 1,4-Dichlorobenzene	12.09	146	494200	28.892	ug/l	99
91) 1,2-Dichlorobenzene	12.39	146	2594904	154.623	ug/l	100
93) 1,2,4-Trichlorobenzene	13.64	180	18324	1.565	ug/l	97
96) 1,2,3-Trichlorobenzene	14.01	180	13726	1.169	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110719\
Data File : VX013349.D
Acq On : 07 Nov 2019 16:16
Operator : JC/SP
Sample : K5626-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS003RW114R-191030

Quant Time: Nov 07 16:52:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

