

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110720\
 Data File : VX019354.D
 Acq On : 06 Nov 2020 19:50
 Operator : JC/MD
 Sample : PB132855ZHE#17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB132855ZHE#17


Quant Time: Nov 07 01:09:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	279603	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	496535	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	470944	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	230723	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	200571	55.62	ug/l	0.00
Spiked Amount			Recovery	=	111.24%	
35) Dibromofluoromethane	5.46	113	156430	50.16	ug/l	0.00
Spiked Amount			Recovery	=	100.32%	
50) Toluene-d8	8.70	98	608970	51.80	ug/l	0.00
Spiked Amount			Recovery	=	103.60%	
62) 4-Bromofluorobenzene	11.12	95	235144	52.81	ug/l	0.00
Spiked Amount			Recovery	=	105.62%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.00	59	1586	2.227	ug/l #	72
13) Acrolein	2.28	56	63	5.841	ug/l #	1
16) Acetone	2.42	43	9604	7.721	ug/l	100
17) Carbon Disulfide	2.55	76	300	0.798	ug/l #	75
18) Methyl Acetate	2.75	43	3613	1.194	ug/l #	85
20) Methylene Chloride	2.83	84	5826	1.866	ug/l	93
25) 2-Butanone	4.65	43	1858	0.948	ug/l	91
29) Tetrahydrofuran	5.11	42	938	0.759	ug/l #	68
31) Cyclohexane	5.63	56	6011	1.319	ug/l #	2
36) 1,1-Dichloropropene	5.63	75	22896	4.994	ug/l #	50
38) Carbon Tetrachloride	5.63	117	29149	6.237	ug/l #	16
43) Isopropyl Acetate	6.42	43	80077	11.234	ug/l	99
48) Methyl methacrylate	7.65	41	4143	1.208	ug/l #	9
49) 1,4-Dioxane	7.84	88	252	3.303	ug/l #	1
51) 4-Methyl-2-Pentanone	8.59	43	78550	19.173	ug/l #	47
54) cis-1,3-Dichloropropene	8.59	75	38770	6.955	ug/l #	59
56) Ethyl methacrylate	9.29	69	1251	0.250	ug/l #	1
57) 1,3-Dichloropropane	9.52	76	6782	1.216	ug/l #	43
59) 2-Hexanone	9.42	43	75855	24.323	ug/l #	20
60) Dibromochloromethane	9.32	129	935	0.260	ug/l #	11
64) Tetrachloroethene	9.32	164	961	0.268	ug/l #	82
74) N-amyl acetate	10.80	43	3130	0.524	ug/l #	44
76) 1,2,3-Trichloropropane	11.12	75	121176	25.436	ug/l #	37
80) 1,3,5-Trimethylbenzene	11.49	105	2952	0.214	ug/l	96
81) trans-1,4-Dichloro-2-buten	11.12	75	121176	66.973	ug/l #	12

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.00 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50

Instrument :

MSVOA_X

ClientSampleId :


PB132855ZHE#17

Tot Ion:168 Resp: 279603

Ion Ratio Lower Upper




168 100

99 62.8 46.6 69.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#11

Tert butyl alcohol

Concen: 2.227 ug/l

RT: 3.00 min Scan# 314

Delta R.T. -0.02 min

Lab File: VX019354.D


Acq: 06 Nov 2020 19:50

Tgt Ion: 59 Resp: 1586

Ion Ratio Lower Upper



59 100


57 0.0 8.4 12.6#



Abundance Ion 59.00 (58.70 to 59.70): VX0

Ion 57.00 (56.70 to 57.70): VX0





#13

Acrolein

Concen: 5.841 ug/l

RT: 2.28 min Scan# 195

Delta R.T. 0.01 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50

Instrument :

MSVOA_X

ClientSampleId :


PB132855ZHE#17

Tot Ion: 56 Resp: 63

Ion Ratio Lower Upper




56 100

55 220.6 54.7 82.1#



Abundance Ion 56.00 (55.70 to 56.70): VX0

Ion 55.00 (54.70 to 55.70): VX0



#16

Acetone

Concen: 7.721 ug/l

RT: 2.42 min Scan# 218

Delta R.T. -0.01 min

Lab File: VX019354.D


Acq: 06 Nov 2020 19:50

Tgt Ion: 43 Resp: 9604

Ion Ratio Lower Upper



43 100


58 33.0 26.3 39.5



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0





#17

Carbon Disulfide

Concen: 0.798 ug/l

RT: 2.55 min Scan# 240

Delta R.T. 0.00 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50

Instrument :

MSVOA_X

ClientSampleId :


PB132855ZHE#17

Tot Ion: 76 Resp: 300

Ion Ratio Lower Upper


76 100

78 0.0 7.3 10.9#





Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0



Time--> 2.50 2.55 2.60



#18

Methyl Acetate

Concen: 1.194 ug/l

RT: 2.75 min Scan# 273

Delta R.T. 0.01 min

Lab File: VX019354.D


Acq: 06 Nov 2020 19:50

Tgt Ion: 43 Resp: 3613

Ion Ratio Lower Upper


43 100

74 20.1 22.6 33.8#





Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 74.00 (73.70 to 74.70): VX0



Time--> 2.70 2.75 2.80









#20
Methvlene Chloride
Concen: 1.866 ug/l
RT: 2.83 min Scan# 286
Delta R.T. -0.01 min
Lab File: VX019354.D
Acq: 06 Nov 2020 19:50




Instrument :
MSVOA_X
ClientSampleId :
PB132855ZHE#17


Tot Ion: 84 Resp: 5826
Ion Ratio Lower Upper
84 100
49 121.2 86.8 130.2
51 34.0 27.3 40.9
86 61.3 50.6 76.0



#25
2-Butanone
Concen: 0.948 ug/l
RT: 4.65 min Scan# 584
Delta R.T. 0.01 min
Lab File: VX019354.D
Acq: 06 Nov 2020 19:50

Tgt Ion: 43 Resp: 1858
Ion Ratio Lower Upper
43 100
72 25.9 24.6 36.8





#29

Tetrahydrofuran

Concen: 0.759 ug/l

RT: 5.11 min Scan# 660

Delta R.T. 0.02 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50

Instrument :

MSVOA_X

ClientSampleId :

PB132855ZHE#17


Tot Ion: 42 Resp: 938

Ion Ratio Lower Upper

42 100

72 40.8 42.6 63.8#


71 16.8 39.7 59.5#





Abundance Ion 42.00 (41.70 to 42.70): VX0

Ion 72.00 (71.70 to 72.70): VX0

Ion 71.00 (70.70 to 71.70): VX0



Time--> 5.00 5.05 5.10 5.15 5.20



#31

Cyclohexane

Concen: 1.319 ug/l

RT: 5.63 min Scan# 745

Delta R.T. 0.08 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50


Tgt Ion: 56 Resp: 6011

Ion Ratio Lower Upper

56 100

69 168.8 25.9 38.9#


84 136.1 73.4 110.2#




Abundance Ion 56.00 (55.70 to 56.70): VX0


Ion 69.00 (68.70 to 69.70): VX0

Ion 84.00 (83.70 to 84.70): VX0



Time--> 5.55 5.60 5.65 5.70 5.75





#33

1,2-Dichloroethane-d4

Concen: 55.624 ug/l

RT: 6.03 min Scan# 811

Delta R.T. -0.00 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50

Instrument :

MSVOA_X

ClientSampleId :


PB132855ZHE#17

Tot Ion: 65 Resp: 200571

Ion Ratio Lower Upper




65 100

67 51.3 0.0 103.0



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 942

Delta R.T. -0.00 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50


Tgt Ion: 114 Resp: 496535

Ion Ratio Lower Upper

114 100

63 21.0 0.0 38.8



88 16.4 0.0 34.2




Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0





#35

Dibromofluoromethane

Concen: 50.155 ug/l

RT: 5.46 min Scan# 718

Delta R.T. -0.00 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50

Instrument :

MSVOA_X

ClientSampleId :

PB132855ZHE#17


Tot Ion:113 Resp: 156430

Ion Ratio Lower Upper

113 100

111 103.3 82.3 123.5




192 19.3 16.2 24.2



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 4.994 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.14 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50


Tgt Ion: 75 Resp: 22896

Ion Ratio Lower Upper

75 100

110 9.9 18.3 54.9#



77 0.0 24.9 37.3#




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VX0





#38

Carbon Tetrachloride

Concen: 6.237 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.12 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50

Instrument :

MSVOA_X

ClientSampleId :

PB132855ZHE#17


Tot Ion:117 Resp: 29149

Ion Ratio Lower Upper

117 100

119 3.7 75.0 112.6#




121 0.0 23.9 35.9#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 11.234 ug/l

RT: 6.42 min Scan# 874

Delta R.T. 0.01 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50


Tgt Ion: 43 Resp: 80077

Ion Ratio Lower Upper

43 100

61 21.6 17.8 26.6



87 15.5 12.6 19.0




Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0









#48
Methvl methacrvlate
Concen: 1.208 ug/l
RT: 7.65 min Scan# 1077
Delta R.T. -0.09 min
Lab File: VX019354.D
Acq: 06 Nov 2020 19:50




Instrument :
MSVOA_X
ClientSampleId :
PB132855ZHE#17


Tot Ion: 41 Resp: 4143
Ion Ratio Lower Upper
41 100
69 0.0 75.5 113.3#
39 0.0 48.9 73.3#



#49
1,4-Dioxane
Concen: 3.303 ug/l
RT: 7.84 min Scan# 1108
Delta R.T. 0.12 min
Lab File: VX019354.D
Acq: 06 Nov 2020 19:50

Tgt Ion: 88 Resp: 252
Ion Ratio Lower Upper
88 100
43 533920.6 23.0 34.6#
58 2363.5 52.2 78.2#





#50

Toluene-d8

Concen: 51.795 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50

Instrument :

MSVOA_X

ClientSampleId :


PB132855ZHE#17

Tot Ion: 98 Resp: 608970

Ion Ratio Lower Upper




98 100

100 65.4 53.0 79.6



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX0



#51

4-Methyl-2-Pentanone

Concen: 19.173 ug/l

RT: 8.59 min Scan# 1231

Delta R.T. -0.03 min

Lab File: VX019354.D


Acq: 06 Nov 2020 19:50

Tgt Ion: 43 Resp: 78550

Ion Ratio Lower Upper



43 100


58 9.4 35.0 52.4#



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0









#54

cis-1,3-Dichloropropene
Concen: 6.955 ug/l
RT: 8.59 min Scan# 1231
Delta R.T. 0.18 min
Lab File: VX019354.D
Acq: 06 Nov 2020 19:50

Instrument :
MSVOA_X
ClientSampleId :
PB132855ZHE#17




Tot Ion: 75 Resp: 38770
Ion Ratio Lower Upper
75 100
77 0.6 25.5 38.3#
39 25.6 36.2 54.2#




#56

Ethyl methacrylate
Concen: 0.250 ug/l
RT: 9.29 min Scan# 1345
Delta R.T. 0.13 min
Lab File: VX019354.D
Acq: 06 Nov 2020 19:50

Tgt Ion: 69 Resp: 1251
Ion Ratio Lower Upper
69 100
41 5593.0 46.5 69.7#
39 2260.0 27.6 41.4#









#57
 1,3-Dichloropropane
 Concen: 1.216 ug/l
 RT: 9.52 min Scan# 1384
 Delta R.T. 0.17 min
 Lab File: VX019354.D
 Acq: 06 Nov 2020 19:50




Instrument :
 MSVOA_X
 ClientSampleId :
 PB132855ZHE#17


Tot Ion: 76 Resp: 6782
 Ion Ratio Lower Upper
 76 100
 78 0.7 25.9 38.9#



#59
 2-Hexanone
 Concen: 24.323 ug/l
 RT: 9.42 min Scan# 1367
 Delta R.T. -0.05 min
 Lab File: VX019354.D
 Acq: 06 Nov 2020 19:50

Tgt Ion: 43 Resp: 75855
 Ion Ratio Lower Upper
 43 100
 58 0.0 30.5 91.5#





#60

Dibromochloromethane

Concen: 0.260 ug/l

RT: 9.32 min Scan# 1350

Delta R.T. -0.24 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50

Instrument :

MSVOA_X

ClientSampleId :


PB132855ZHE#17

Tot Ion:129 Resp: 935

Ion Ratio Lower Upper

129 100

127 0.0 38.6 115.8#



Abundance Ion 128.80 (128.50 to 129.50): V

Ion 126.80 (126.50 to 127.50): V

9.32



600

400

200

0

Time-->



#62

4-Bromofluorobenzene

Concen: 52.811 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.00 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50


Tgt Ion: 95 Resp: 235144

Ion Ratio Lower Upper

95 100

174 74.8 0.0 156.0

176 72.8 0.0 149.8



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

250000

200000


150000


100000

50000

0

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.10 min Scan# 1478
Delta R.T. -0.00 min
Lab File: VX019354.D
Acq: 06 Nov 2020 19:50




Instrument :
MSVOA_X
ClientSampleId :
PB132855ZHE#17


Tot Ion:117 Resp: 470944
Ion Ratio Lower Upper
117 100
82 57.2 43.6 65.4
119 32.3 25.5 38.3



#64
Tetrachloroethene
Concen: 0.268 ug/l
RT: 9.32 min Scan# 1350
Delta R.T. -0.00 min
Lab File: VX019354.D
Acq: 06 Nov 2020 19:50

Tgt Ion:164 Resp: 961
Ion Ratio Lower Upper
164 100
166 104.7 104.1 156.1
129 86.8 76.2 114.4
131 73.5 74.6 112.0#





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.06 min Scan# 1800

Delta R.T. -0.00 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50

Instrument :

MSVOA_X

ClientSampleId :

PB132855ZHE#17


Tot Ion:152 Resp: 230723

Ion Ratio Lower Upper

152 100

115 60.9 41.9 125.7




150 157.0 0.0 347.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 0.524 ug/l

RT: 10.80 min Scan# 1594

Delta R.T. -0.07 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50

Tgt Ion: 43 Resp: 3130


Ion Ratio Lower Upper

43 100

70 0.0 39.6 59.4#

55 10.4 20.7 31.1#

61 0.0 21.9 32.9#






Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0









#76
 1,2,3-Trichloropropane
 Concen: 25.436 ug/l
 RT: 11.12 min Scan# 1646
 Delta R.T. -0.16 min
 Lab File: VX019354.D
 Acq: 06 Nov 2020 19:50




Instrument :
 MSVOA_X
 ClientSampleId :
 PB132855ZHE#17


Tot Ion: 75 Resp: 121176
 Ion Ratio Lower Upper
 75 100
 77 1.2 20.3 60.9#



#80
 1,3,5-Trimethylbenzene
 Concen: 0.214 ug/l
 RT: 11.49 min Scan# 1707
 Delta R.T. 0.01 min
 Lab File: VX019354.D
 Acq: 06 Nov 2020 19:50

Tgt Ion: 105 Resp: 2952
 Ion Ratio Lower Upper
 105 100
 120 45.9 24.3 72.8





#81

trans-1,4-Dichloro-2-butene

Concen: 66.973 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. 0.07 min

Lab File: VX019354.D

Acq: 06 Nov 2020 19:50

Instrument :

MSVOA_X

ClientSampleId :

PB132855ZHE#17

Tot Ion: 75 Resp: 121176

Ion Ratio Lower Upper

75 100

53 0.0 73.3 109.9#

89 0.0 38.2 57.2#

