

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110720\
 Data File : VX019336.D
 Acq On : 06 Nov 2020 13:06
 Operator : JC/MD
 Sample : L4616-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA20-GMZ5-RB1

Quant Time: Nov 07 01:53:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	291393	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	515750	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	483461	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	244082	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	202802	53.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.94%	
35) Dibromofluoromethane	5.46	113	159470	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
50) Toluene-d8	8.70	98	623504	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
62) 4-Bromofluorobenzene	11.12	95	243197	52.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.16%	
Target Compounds						
3) Chloromethane	1.31	50	1320	0.486	ug/l	Qvalue 100
16) Acetone	2.42	43	10826	8.351	ug/l	98
20) Methylene Chloride	2.84	84	1357	0.417	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110720\
Data File : VX019336.D
Acq On : 06 Nov 2020 13:06
Operator : JC/MD
Sample : L4616-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
FA20-GMZ5-RB1

Quant Time: Nov 07 01:53:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

