

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110720\
 Data File : VX019350.D
 Acq On : 06 Nov 2020 18:20
 Operator : JC/MD
 Sample : L4666-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 07 02:52:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	288658	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	502392	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	476400	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	237423	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	204069	54.82	ug/l	0.00
Spiked Amount						
			Recovery	=	109.64%	
35) Dibromofluoromethane	5.46	113	157225	49.82	ug/l	0.00
Spiked Amount						
			Recovery	=	99.64%	
50) Toluene-d8	8.70	98	620718	52.18	ug/l	0.00
Spiked Amount						
			Recovery	=	104.36%	
62) 4-Bromofluorobenzene	11.12	95	239948	53.26	ug/l	0.00
Spiked Amount						
			Recovery	=	106.52%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110720\
Data File : VX019350.D
Acq On : 06 Nov 2020 18:20
Operator : JC/MD
Sample : L4666-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 07 02:52:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

