

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110720\
 Data File : VX019354.D
 Acq On : 06 Nov 2020 19:50
 Operator : JC/MD
 Sample : PB132855ZHE#17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB132855ZHE#17

Quant Time: Nov 07 03:08:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	279603	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	496535	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	470944	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	230723	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	200571	55.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.24%	
35) Dibromofluoromethane	5.46	113	156430	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
50) Toluene-d8	8.70	98	608970	51.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.60%	
62) 4-Bromofluorobenzene	11.12	95	235144	52.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.62%	
Target Compounds						
16) Acetone	2.42	43	9604	7.721	ug/l	100
18) Methyl Acetate	2.75	43	3613	1.194	ug/l #	85
20) Methylene Chloride	2.83	84	5826	1.866	ug/l	93
43) Isopropyl Acetate	6.42	43	80077	11.234	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110720\
Data File : VX019354.D
Acq On : 06 Nov 2020 19:50
Operator : JC/MD
Sample : PB132855ZHE#17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB132855ZHE#17

Quant Time: Nov 07 03:08:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

