

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110720\
 Data File : VX019329.D
 Acq On : 06 Nov 2020 10:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 07 01:32:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.417	0.425	-1.9	93	0.00
3 P	Chloromethane	0.466	0.458	1.7	93	0.00
4 C	Vinyl Chloride	0.540	0.565	-4.6#	95	0.00
5 T	Bromomethane	0.369	0.360	2.4	90	-0.01
6 T	Chloroethane	0.345	0.374	-8.4	97	0.00
7 T	Trichlorofluoromethane	0.848	0.890	-5.0	93	0.00
8 T	Diethyl Ether	0.326	0.354	-8.6	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.470	0.505	-7.4	96	0.00
10 T	Methyl Iodide	0.478	0.523	-9.4	91	0.00
11 T	Tert butyl alcohol	0.127	0.131	-3.1	95	-0.01
12 CM	1,1-Dichloroethene	0.485	0.504	-3.9#	96	0.00
13 T	Acrolein	0.062	0.058	6.5	97	0.00
14 T	Allyl chloride	0.734	0.816	-11.2	100	0.00
15 T	Acrylonitrile	0.268	0.293	-9.3	96	0.00
16 T	Acetone	0.222	0.257	-15.8	105	0.00
17 T	Carbon Disulfide	1.473	1.339	9.1	90	0.00
18 T	Methyl Acetate	0.541	0.600	-10.9	101	0.00
19 T	Methyl tert-butyl Ether	1.679	1.842	-9.7	97	0.00
20 T	Methylene Chloride	0.558	0.577	-3.4	96	0.00
21 T	trans-1,2-Dichloroethene	0.558	0.568	-1.8	94	0.00
22 T	Diisopropyl ether	1.463	1.681	-14.9	101	0.00
23 T	Vinyl Acetate	1.305	1.477	-13.2	98	0.00
24 P	1,1-Dichloroethane	0.914	1.002	-9.6	97	0.00
25 T	2-Butanone	0.351	0.395	-12.5	99	-0.01
26 T	2,2-Dichloropropane	0.859	0.944	-9.9	98	0.00
27 T	cis-1,2-Dichloroethene	0.623	0.650	-4.3	94	0.00
28 T	Bromochloromethane	0.415	0.415	0.0	97	0.00
29 T	Tetrahydrofuran	0.221	0.246	-11.3	95	0.00
30 C	Chloroform	0.973	1.058	-8.7#	96	0.00
31 T	Cyclohexane	0.815	0.871	-6.9	97	0.00
32 T	1,1,1-Trichloroethane	0.882	0.950	-7.7	95	0.00
33 S	1,2-Dichloroethane-d4	0.645	0.615	4.7	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.314	0.297	5.4	93	0.00
36 T	1,1-Dichloropropene	0.462	0.491	-6.3	97	0.00
37 T	Ethyl Acetate	0.429	0.462	-7.7	95	-0.01
38 T	Carbon Tetrachloride	0.471	0.497	-5.5	93	0.00
39 T	Methylcyclohexane	0.552	0.579	-4.9	93	0.00
40 TM	Benzene	1.309	1.397	-6.7	95	0.00
41 T	Methacrylonitrile	0.231	0.247	-6.9	95	0.00
42 TM	1,2-Dichloroethane	0.482	0.519	-7.7	96	0.00
43 T	Isopropyl Acetate	0.718	0.776	-8.1	96	0.00
44 TM	Trichloroethene	0.375	0.385	-2.7	94	0.00
45 C	1,2-Dichloropropane	0.321	0.356	-10.9#	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX110720\
 Data File : VX019329.D
 Acq On : 06 Nov 2020 10:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 07 01:32:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.237	0.249	-5.1	95	0.00
47 T	Bromodichloromethane	0.467	0.514	-10.1	96	0.00
48 T	Methyl methacrylate	0.345	0.380	-10.1	96	0.00
49 T	1,4-Dioxane	0.008	0.008	0.0	89	-0.01
50 S	Toluene-d8	1.184	1.112	6.1	92	0.00
51 T	4-Methyl-2-Pentanone	0.413	0.461	-11.6	96	0.00
52 CM	Toluene	0.843	0.908	-7.7#	96	0.00
53 T	t-1,3-Dichloropropene	0.534	0.582	-9.0	97	0.00
54 T	cis-1,3-Dichloropropene	0.561	0.619	-10.3	97	0.00
55 T	1,1,2-Trichloroethane	0.333	0.358	-7.5	96	0.00
56 T	Ethyl methacrylate	0.503	0.564	-12.1	96	0.00
57 T	1,3-Dichloropropane	0.562	0.611	-8.7	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.281	0.293	-4.3	95	0.00
59 T	2-Hexanone	0.314	0.362	-15.3	99	0.00
60 T	Dibromochloromethane	0.362	0.387	-6.9	92	0.00
61 T	1,2-Dibromoethane	0.356	0.381	-7.0	95	0.00
62 S	4-Bromofluorobenzene	0.448	0.448	0.0	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.381	0.398	-4.5	98	0.00
65 PM	Chlorobenzene	1.022	1.054	-3.1	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.368	0.382	-3.8	93	0.00
67 C	Ethyl Benzene	1.821	1.938	-6.4#	97	0.00
68 T	m/p-Xylenes	0.683	0.727	-6.4	96	0.00
69 T	o-Xylene	0.648	0.691	-6.6	96	0.00
70 T	Styrene	1.106	1.204	-8.9	98	0.00
71 P	Bromoform	0.292	0.306	-4.8	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.539	3.616	-2.2	98	0.00
74 T	N-amyl acetate	1.294	1.394	-7.7	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.106	1.159	-4.8	100	0.00
76 T	1,2,3-Trichloropropane	1.032	1.130	-9.5	104	0.00
77 T	Bromobenzene	0.878	0.873	0.6	95	0.00
78 T	n-propylbenzene	4.040	4.215	-4.3	100	0.00
79 T	2-Chlorotoluene	2.379	2.489	-4.6	101	0.00
80 T	1,3,5-Trimethylbenzene	2.990	3.107	-3.9	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.392	0.396	-1.0	97	0.00
82 T	4-Chlorotoluene	2.905	2.978	-2.5	101	0.00
83 T	tert-Butylbenzene	2.863	2.914	-1.8	97	0.00
84 T	1,2,4-Trimethylbenzene	3.019	3.121	-3.4	99	0.00
85 T	sec-Butylbenzene	3.434	3.536	-3.0	98	0.00
86 T	p-Isopropyltoluene	3.169	3.286	-3.7	98	0.00
87 T	1,3-Dichlorobenzene	1.636	1.667	-1.9	100	0.00
88 T	1,4-Dichlorobenzene	1.703	1.701	0.1	102	0.00
89 T	n-Butylbenzene	2.979	3.103	-4.2	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110720\
Data File : VX019329.D
Acq On : 06 Nov 2020 10:02
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 07 01:32:19 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.539	0.565	-4.8	97	0.00
91 T	1,2-Dichlorobenzene	1.571	1.590	-1.2	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.271	0.266	1.8	93	0.00
93 T	1,2,4-Trichlorobenzene	1.180	1.083	8.2	92	0.00
94 T	Hexachlorobutadiene	0.519	0.450	13.3	89	0.00
95 T	Naphthalene	3.716	3.579	3.7	95	0.00
96 T	1,2,3-Trichlorobenzene	1.145	1.046	8.6	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6