

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110720\
 Data File : VX019355.D
 Acq On : 06 Nov 2020 20:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 07 03:09:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.417	0.426	-2.2	83	0.00
3 P	Chloromethane	0.466	0.502	-7.7	90	0.00
4 C	Vinyl Chloride	0.540	0.596	-10.4#	89	0.00
5 T	Bromomethane	0.369	0.407	-10.3	90	-0.01
6 T	Chloroethane	0.345	0.401	-16.2	92	0.00
7 T	Trichlorofluoromethane	0.848	0.921	-8.6	85	0.00
8 T	Diethyl Ether	0.326	0.378	-16.0	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.470	0.496	-5.5	84	0.00
10 T	Methyl Iodide	0.478	0.480	-0.4	74	0.00
11 T	Tert butyl alcohol	0.127	0.153	-20.5	98	-0.01
12 CM	1,1-Dichloroethene	0.485	0.515	-6.2#	86	0.00
13 T	Acrolein	0.062	0.059	4.8	86	0.00
14 T	Allyl chloride	0.734	0.845	-15.1	92	0.00
15 T	Acrylonitrile	0.268	0.322	-20.1	94	0.00
16 T	Acetone	0.222	0.282	-27.0#	102	0.00
17 T	Carbon Disulfide	1.473	1.324	10.1	79	0.00
18 T	Methyl Acetate	0.541	0.796	-47.1#	119	0.00
19 T	Methyl tert-butyl Ether	1.679	1.949	-16.1	91	0.00
20 T	Methylene Chloride	0.558	0.620	-11.1	91	0.00
21 T	trans-1,2-Dichloroethene	0.558	0.587	-5.2	86	0.00
22 T	Diisopropyl ether	1.463	1.782	-21.8	95	0.00
23 T	Vinyl Acetate	1.305	1.512	-15.9	89	0.00
24 P	1,1-Dichloroethane	0.914	1.070	-17.1	92	0.00
25 T	2-Butanone	0.351	0.433	-23.4	96	0.00
26 T	2,2-Dichloropropane	0.859	0.866	-0.8	80	0.00
27 T	cis-1,2-Dichloroethene	0.623	0.683	-9.6	88	0.00
28 T	Bromochloromethane	0.415	0.464	-11.8	96	0.00
29 T	Tetrahydrofuran	0.221	0.275	-24.4	94	0.00
30 C	Chloroform	0.973	1.127	-15.8#	90	0.00
31 T	Cyclohexane	0.815	0.896	-9.9	88	0.00
32 T	1,1,1-Trichloroethane	0.882	1.001	-13.5	89	0.00
33 S	1,2-Dichloroethane-d4	0.645	0.694	-7.6	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.314	0.316	-0.6	92	0.00
36 T	1,1-Dichloropropene	0.462	0.475	-2.8	88	0.00
37 T	Ethyl Acetate	0.429	0.480	-11.9	92	0.00
38 T	Carbon Tetrachloride	0.471	0.487	-3.4	85	0.00
39 T	Methylcyclohexane	0.552	0.544	1.4	82	0.00
40 TM	Benzene	1.309	1.420	-8.5	90	0.00
41 T	Methacrylonitrile	0.231	0.267	-15.6	95	0.00
42 TM	1,2-Dichloroethane	0.482	0.531	-10.2	92	0.00
43 T	Isopropyl Acetate	0.718	0.793	-10.4	92	0.00
44 TM	Trichloroethene	0.375	0.373	0.5	85	0.00
45 C	1,2-Dichloropropane	0.321	0.361	-12.5#	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX110720\
 Data File : VX019355.D
 Acq On : 06 Nov 2020 20:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 07 03:09:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.237	0.249	-5.1	88	0.00
47 T	Bromodichloromethane	0.467	0.520	-11.3	90	0.00
48 T	Methyl methacrylate	0.345	0.399	-15.7	94	0.00
49 T	1,4-Dioxane	0.008	0.009	-12.5	96	-0.01
50 S	Toluene-d8	1.184	1.188	-0.3	91	0.00
51 T	4-Methyl-2-Pentanone	0.413	0.495	-19.9	96	0.00
52 CM	Toluene	0.843	0.913	-8.3#	90	0.00
53 T	t-1,3-Dichloropropene	0.534	0.572	-7.1	89	0.00
54 T	cis-1,3-Dichloropropene	0.561	0.605	-7.8	88	0.00
55 T	1,1,2-Trichloroethane	0.333	0.368	-10.5	91	0.00
56 T	Ethyl methacrylate	0.503	0.577	-14.7	91	0.00
57 T	1,3-Dichloropropane	0.562	0.625	-11.2	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.281	0.304	-8.2	92	0.00
59 T	2-Hexanone	0.314	0.378	-20.4	96	0.00
60 T	Dibromochloromethane	0.362	0.396	-9.4	87	0.00
61 T	1,2-Dibromoethane	0.356	0.389	-9.3	91	0.00
62 S	4-Bromofluorobenzene	0.448	0.476	-6.2	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.381	0.357	6.3	84	0.00
65 PM	Chlorobenzene	1.022	1.034	-1.2	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.368	0.380	-3.3	89	0.00
67 C	Ethyl Benzene	1.821	1.887	-3.6#	91	0.00
68 T	m/p-Xylenes	0.683	0.709	-3.8	90	0.00
69 T	o-Xylene	0.648	0.676	-4.3	91	0.00
70 T	Styrene	1.106	1.174	-6.1	92	0.00
71 P	Bromoform	0.292	0.306	-4.8	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.539	3.572	-0.9	91	0.00
74 T	N-amyl acetate	1.294	1.441	-11.4	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.106	1.220	-10.3	99	0.00
76 T	1,2,3-Trichloropropane	1.032	1.112	-7.8	96	0.00
77 T	Bromobenzene	0.878	0.845	3.8	87	0.00
78 T	n-propylbenzene	4.040	4.138	-2.4	92	0.00
79 T	2-Chlorotoluene	2.379	2.455	-3.2	94	0.00
80 T	1,3,5-Trimethylbenzene	2.990	3.042	-1.7	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.392	0.386	1.5	89	0.00
82 T	4-Chlorotoluene	2.905	2.947	-1.4	94	0.00
83 T	tert-Butylbenzene	2.863	2.945	-2.9	92	0.00
84 T	1,2,4-Trimethylbenzene	3.019	3.090	-2.4	92	0.00
85 T	sec-Butylbenzene	3.434	3.491	-1.7	91	0.00
86 T	p-Isopropyltoluene	3.169	3.191	-0.7	89	0.00
87 T	1,3-Dichlorobenzene	1.636	1.621	0.9	92	0.00
88 T	1,4-Dichlorobenzene	1.703	1.632	4.2	92	0.00
89 T	n-Butylbenzene	2.979	2.940	1.3	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110720\
Data File : VX019355.D
Acq On : 06 Nov 2020 20:13
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 07 03:09:29 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.539	0.558	-3.5	90	0.00
91 T	1,2-Dichlorobenzene	1.571	1.585	-0.9	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.271	0.276	-1.8	91	0.00
93 T	1,2,4-Trichlorobenzene	1.180	1.053	10.8	84	0.00
94 T	Hexachlorobutadiene	0.519	0.408	21.4	76	0.00
95 T	Naphthalene	3.716	3.566	4.0	90	0.00
96 T	1,2,3-Trichlorobenzene	1.145	1.008	12.0	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6