

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX110720\
 Data File : VX019352.D
 Acq On : 06 Nov 2020 19:05
 Operator : JC/MD
 Sample : L4666-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Nov 07 03:00:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	283190	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	500018	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	468488	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	227940	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	201199	55.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.18%	
35) Dibromofluoromethane	5.46	113	155367	49.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.94%	
50) Toluene-d8	8.70	98	606186	51.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.40%	
62) 4-Bromofluorobenzene	11.12	95	236647	52.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.56%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX110720\
Data File : VX019352.D
Acq On : 06 Nov 2020 19:05
Operator : JC/MD
Sample : L4666-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Nov 07 03:00:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. 0.00 min

Lab File: VX019352.D

Acq: 06 Nov 2020 19:05

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-SAMPLE-MANA

Tot Ion:168 Resp: 283190

Ion Ratio Lower Upper

168 100

99 61.9 46.6 69.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.63

100000

80000

60000



40000

20000

0

Time-->

5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 55.092 ug/l

RT: 6.03 min Scan# 811

Delta R.T. 0.00 min

Lab File: VX019352.D


Acq: 06 Nov 2020 19:05

Tgt Ion: 65 Resp: 201199

Ion Ratio Lower Upper

65 100

67 51.5 0.0 103.0



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.03

80000

60000


40000


20000

0

Time-->

5.90 6.00 6.10 6.20 6.30





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 942

Delta R.T. 0.00 min

Lab File: VX019352.D

Acq: 06 Nov 2020 19:05

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-SAMPLE-MANA


Tot Ion:114 Resp: 500018

Ion Ratio Lower Upper

114 100

63 21.0 0.0 38.8




88 16.4 0.0 34.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 49.467 ug/l

RT: 5.46 min Scan# 718

Delta R.T. 0.00 min

Lab File: VX019352.D

Acq: 06 Nov 2020 19:05


Tgt Ion:113 Resp: 155367

Ion Ratio Lower Upper

113 100

111 102.9 82.3 123.5



192 19.1 16.2 24.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 51.199 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. 0.00 min

Lab File: VX019352.D

Acq: 06 Nov 2020 19:05

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-SAMPLE-MANA

Tot Ion: 98 Resp: 606186

Ion Ratio Lower Upper




98 100

100 65.9 53.0 79.6



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



#62

4-Bromofluorobenzene

Concen: 52.779 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. 0.00 min

Lab File: VX019352.D

Acq: 06 Nov 2020 19:05


Tgt Ion: 95 Resp: 236647

Ion Ratio Lower Upper

95 100

174 73.6 0.0 156.0



176 71.9 0.0 149.8




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.10 min Scan# 1478
Delta R.T. 0.00 min
Lab File: VX019352.D
Acq: 06 Nov 2020 19:05

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Tot Ion:117 Resp: 468488
Ion Ratio Lower Upper
117 100
82 56.5 43.6 65.4
119 32.8 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.06 min Scan# 1800
Delta R.T. 0.00 min
Lab File: VX019352.D
Acq: 06 Nov 2020 19:05

Tgt Ion:152 Resp: 227940
Ion Ratio Lower Upper
152 100
115 61.6 41.9 125.7
150 160.2 0.0 347.8

