

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110722\
 Data File : VX032616.D
 Acq On : 07 Nov 2022 10:55
 Operator : JC/MD
 Sample : VX1107MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1107MBL01

Quant Time: Nov 08 03:09:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	119333	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	195831	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	173346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83054	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83271	50.221	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.440%
35) Dibromofluoromethane	5.385	113	64793	46.384	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.760%
50) Toluene-d8	8.647	98	235155	48.413	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.820%
62) 4-Bromofluorobenzene	11.079	95	92562	48.785	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110722\
Data File : VX032616.D
Acq On : 07 Nov 2022 10:55
Operator : JC/MD
Sample : VX1107MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1107MBL01

Quant Time: Nov 08 03:09:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

