

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110722\
 Data File : VX032625.D
 Acq On : 07 Nov 2022 14:28
 Operator : JC/MD
 Sample : N5484-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-17

Quant Time: Nov 08 03:13:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	103841	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	171147	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	150854	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69618	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72140	49.998	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.000%
35) Dibromofluoromethane	5.391	113	56887	46.598	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.200%
50) Toluene-d8	8.653	98	205756	48.470	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.940%
62) 4-Bromofluorobenzene	11.079	95	77032	46.455	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.920%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	1040	1.983	ug/l	97
20) Methylene Chloride	2.788	84	613	0.510	ug/l	93
30) Chloroform	5.099	83	2854	1.259	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110722\
Data File : VX032625.D
Acq On : 07 Nov 2022 14:28
Operator : JC/MD
Sample : N5484-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-17

Quant Time: Nov 08 03:13:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

