

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
 Data File : VX043727.D
 Acq On : 07 Nov 2024 11:29
 Operator : JC/MD
 Sample : VX1107MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1107MBL01

Quant Time: Nov 08 04:03:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	104295	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	185184	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	164618	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72720	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	63523	38.198	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	76.400%
35) Dibromofluoromethane	5.379	113	59687	47.532	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.060%
50) Toluene-d8	8.646	98	212673	48.926	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.860%
62) 4-Bromofluorobenzene	11.079	95	74354	46.162	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
Data File : VX043727.D
Acq On : 07 Nov 2024 11:29
Operator : JC/MD
Sample : VX1107MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1107MBL01

Quant Time: Nov 08 04:03:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

