

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
 Data File : VX043728.D
 Acq On : 07 Nov 2024 11:53
 Operator : JC/MD
 Sample : VX1107WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1107WBL01

Quant Time: Nov 08 04:04:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	93207	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	162976	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	144247	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66166	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	56688	38.143	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	76.280%
35) Dibromofluoromethane	5.385	113	50670	45.850	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.700%
50) Toluene-d8	8.647	98	186832	48.838	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.680%
62) 4-Bromofluorobenzene	11.079	95	66258	46.741	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
Data File : VX043728.D
Acq On : 07 Nov 2024 11:53
Operator : JC/MD
Sample : VX1107WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1107WBL01

Quant Time: Nov 08 04:04:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

