

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
 Data File : VX043731.D
 Acq On : 07 Nov 2024 13:10
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Nov 08 04:06:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	104674	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	189853	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	169285	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	72933	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	67107	40.207	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	80.420%
35) Dibromofluoromethane	5.385	113	58541	45.473	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.940%
50) Toluene-d8	8.647	98	223335	50.115	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.220%
62) 4-Bromofluorobenzene	11.079	95	76629	46.404	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
Data File : VX043731.D
Acq On : 07 Nov 2024 13:10
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Nov 08 04:06:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

