

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
 Data File : VX043732.D
 Acq On : 07 Nov 2024 13:33
 Operator : JC/MD
 Sample : PB164695TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB164695TB

Quant Time: Nov 08 04:06:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	73611	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	130286	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119649	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49489	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	42900	36.550	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	73.100%#
35) Dibromofluoromethane	5.385	113	39125	44.286	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.580%
50) Toluene-d8	8.647	98	154736	50.597	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.200%
62) 4-Bromofluorobenzene	11.079	95	51966	45.856	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.720%
Target Compounds						
						Qvalue
16) Acetone	2.398	43	11984	25.711	ug/l	95
20) Methylene Chloride	2.788	84	2152	2.142	ug/l #	80
43) Isopropyl Acetate	6.354	43	44573	21.192	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
Data File : VX043732.D
Acq On : 07 Nov 2024 13:33
Operator : JC/MD
Sample : PB164695TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB164695TB

Quant Time: Nov 08 04:06:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

