

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
 Data File : VX043733.D
 Acq On : 07 Nov 2024 13:56
 Operator : JC/MD
 Sample : P4718-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WB-307-SB02

Quant Time: Nov 08 04:07:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	74982	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	134619	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118244	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50308	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	43950	36.760	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	73.520%#		
35) Dibromofluoromethane	5.385	113	39878	43.685	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	87.380%		
50) Toluene-d8	8.647	98	157069	49.706	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.420%		
62) 4-Bromofluorobenzene	11.079	95	52711	45.017	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	90.040%		
Target Compounds						
16) Acetone	2.404	43	8728	18.383	ug/l	95
20) Methylene Chloride	2.788	84	3524	3.444	ug/l #	87
43) Isopropyl Acetate	6.348	43	39098	17.991	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
Data File : VX043733.D
Acq On : 07 Nov 2024 13:56
Operator : JC/MD
Sample : P4718-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WB-307-SB02

Quant Time: Nov 08 04:07:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

