

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
 Data File : VX043735.D
 Acq On : 07 Nov 2024 14:43
 Operator : JC/MD
 Sample : P4720-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JC-701-COMP-01

Quant Time: Nov 08 04:08:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	74773	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	137189	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121051	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56074	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	41737	35.006	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	70.020%#
35) Dibromofluoromethane	5.385	113	38223	41.088	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	82.180%
50) Toluene-d8	8.647	98	157115	48.789	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.580%
62) 4-Bromofluorobenzene	11.079	95	53267	44.639	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.280%
Target Compounds						
					Qvalue	
16) Acetone	2.410	43	11174	23.601	ug/l	92
18) Methyl Acetate	2.721	43	1621	1.171	ug/l #	90
20) Methylene Chloride	2.794	84	1914	1.876	ug/l #	72
43) Isopropyl Acetate	6.355	43	27878	12.588	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
Data File : VX043735.D
Acq On : 07 Nov 2024 14:43
Operator : JC/MD
Sample : P4720-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JC-701-COMP-01

Quant Time: Nov 08 04:08:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

