

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
 Data File : VX043737.D
 Acq On : 07 Nov 2024 15:29
 Operator : JC/MD
 Sample : P4722-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-2(4)

Quant Time: Nov 08 04:09:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	72799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	128565	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115248	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49956	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	40061	34.512	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	69.020%#		
35) Dibromofluoromethane	5.385	113	36048	41.349	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	82.700%		
50) Toluene-d8	8.647	98	148610	49.244	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.480%		
62) 4-Bromofluorobenzene	11.079	95	51768	46.293	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	92.580%		
Target Compounds						
16) Acetone	2.410	43	6855	14.871	ug/l	90
18) Methyl Acetate	2.727	43	1832	1.360	ug/l	98
20) Methylene Chloride	2.788	84	1932	1.945	ug/l #	85
43) Isopropyl Acetate	6.355	43	40835	19.675	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
Data File : VX043737.D
Acq On : 07 Nov 2024 15:29
Operator : JC/MD
Sample : P4722-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-2(4)

Quant Time: Nov 08 04:09:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

