

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
 Data File : VX043738.D
 Acq On : 07 Nov 2024 15:52
 Operator : JC/MD
 Sample : P4722-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-3(3)

Quant Time: Nov 08 04:09:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	71929	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	126487	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48006	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	40416	35.239	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	70.480%#
35) Dibromofluoromethane	5.385	113	35599	41.505	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	83.000%
50) Toluene-d8	8.647	98	149442	50.333	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.660%
62) 4-Bromofluorobenzene	11.079	95	48874	44.423	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.840%
Target Compounds						
						Qvalue
16) Acetone	2.404	43	6869	15.082	ug/l	99
20) Methylene Chloride	2.782	84	1807	1.841	ug/l	96
43) Isopropyl Acetate	6.355	43	40535	19.851	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
Data File : VX043738.D
Acq On : 07 Nov 2024 15:52
Operator : JC/MD
Sample : P4722-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-3(3)

Quant Time: Nov 08 04:09:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

