

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
 Data File : VX043740.D
 Acq On : 07 Nov 2024 16:39
 Operator : JC/MD
 Sample : P4722-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-2(2)

Quant Time: Nov 08 04:10:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	70989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129776	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	114600	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50426	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	41199	36.397	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	72.800%#
35) Dibromofluoromethane	5.385	113	36750	41.761	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	83.520%
50) Toluene-d8	8.647	98	149428	49.053	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.100%
62) 4-Bromofluorobenzene	11.085	95	51448	45.578	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.160%
Target Compounds						
					Qvalue	
16) Acetone	2.410	43	6780	15.084	ug/l #	88
18) Methyl Acetate	2.727	43	1931	1.470	ug/l #	76
20) Methylene Chloride	2.788	84	2066	2.133	ug/l #	74
43) Isopropyl Acetate	6.348	43	42473	20.273	ug/l	95
95) Naphthalene	13.780	128	5257	1.339	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
Data File : VX043740.D
Acq On : 07 Nov 2024 16:39
Operator : JC/MD
Sample : P4722-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-2(2)

Quant Time: Nov 08 04:10:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

