

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
 Data File : VX043741.D
 Acq On : 07 Nov 2024 17:02
 Operator : JC/MD
 Sample : P4722-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-3(5)

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 11/08/2024
 Supervised By :Semsettin Yesilyurt 11/08/2024

Quant Time: Nov 08 04:11:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	78752	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	139931	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123692	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53786	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	42738	34.035	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	68.060%	#	
35) Dibromofluoromethane	5.385	113	39285	41.402	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	82.800%		
50) Toluene-d8	8.647	98	163330	49.726	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.460%		
62) 4-Bromofluorobenzene	11.079	95	53343	43.827	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	87.660%		
Target Compounds						
16) Acetone	2.404	43	11210	22.481	ug/l	99
18) Methyl Acetate	2.727	43	1528	1.048	ug/l	97
20) Methylene Chloride	2.788	84	1851m	1.722	ug/l	
43) Isopropyl Acetate	6.355	43	41283	18.275	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
Data File : VX043741.D
Acq On : 07 Nov 2024 17:02
Operator : JC/MD
Sample : P4722-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-3(5)

Quant Time: Nov 08 04:11:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 11/08/2024
Supervised By :Semsettin Yesilyurt 11/08/2024

