

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
 Data File : VX043742.D
 Acq On : 07 Nov 2024 17:25
 Operator : JC/MD
 Sample : P4732-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PPE-Grab

Quant Time: Nov 08 04:11:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	81390	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	147031	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128822	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57793	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	47091	36.286	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	72.580%#
35) Dibromofluoromethane	5.385	113	41568	41.693	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	83.380%
50) Toluene-d8	8.647	98	170800	49.489	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.980%
62) 4-Bromofluorobenzene	11.079	95	57345	44.840	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.680%
Target Compounds						
16) Acetone	2.410	43	8697	16.876	ug/l	96
18) Methyl Acetate	2.727	43	1613	1.071	ug/l	99
20) Methylene Chloride	2.788	84	2246	2.022	ug/l #	84
43) Isopropyl Acetate	6.355	43	47109	19.847	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
Data File : VX043742.D
Acq On : 07 Nov 2024 17:25
Operator : JC/MD
Sample : P4732-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PPE-Grab

Quant Time: Nov 08 04:11:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

