

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
 Data File : VX043726.D
 Acq On : 07 Nov 2024 10:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 08 04:02:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	65	0.00
2 T	Dichlorodifluoromethane	50.000	39.470	21.1	54	0.00
3 P	Chloromethane	50.000	39.104	21.8	53	0.00
4 C	Vinyl Chloride	50.000	40.735	18.5#	53	0.00
5 T	Bromomethane	50.000	47.980	4.0	62	0.00
6 T	Chloroethane	50.000	55.373	-10.7	71	0.00
7 T	Trichlorofluoromethane	50.000	47.030	5.9	61	0.01
8 T	Diethyl Ether	50.000	49.484	1.0	65	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.142	-0.3	65	0.00
10 T	Methyl Iodide	50.000	51.240	-2.5	65	0.00
11 T	Tert butyl alcohol	250.000	263.716	-5.5	66	-0.05
12 CM	1,1-Dichloroethene	50.000	46.183	7.6#	60	0.00
13 T	Acrolein	250.000	200.872	19.7	56	0.00
14 T	Allyl chloride	50.000	45.655	8.7	58	0.00
15 T	Acrylonitrile	250.000	263.537	-5.4	66	-0.01
16 T	Acetone	250.000	255.512	-2.2	67	-0.01
17 T	Carbon Disulfide	50.000	38.822	22.4	50	0.00
18 T	Methyl Acetate	50.000	48.929	2.1	64	0.00
19 T	Methyl tert-butyl Ether	50.000	48.632	2.7	62	-0.01
20 T	Methylene Chloride	50.000	47.841	4.3	64	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.053	7.9	60	0.00
22 T	Diisopropyl ether	50.000	49.015	2.0	63	0.00
23 T	Vinyl Acetate	250.000	248.326	0.7	61	0.00
24 P	1,1-Dichloroethane	50.000	47.492	5.0	61	0.00
25 T	2-Butanone	250.000	258.596	-3.4	64	-0.01
26 T	2,2-Dichloropropane	50.000	46.968	6.1	62	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.224	-2.4	65	0.00
28 T	Bromochloromethane	50.000	48.391	3.2	65	0.00
29 T	Tetrahydrofuran	250.000	247.417	1.0	62	-0.01
30 C	Chloroform	50.000	49.209	1.6#	64	-0.01
31 T	Cyclohexane	50.000	43.034	13.9	56	0.00
32 T	1,1,1-Trichloroethane	50.000	44.917	10.2	57	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.022	8.0	61	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	65	0.00
35 S	Dibromofluoromethane	50.000	52.798	-5.6	69	0.00
36 T	1,1-Dichloropropene	50.000	44.233	11.5	59	0.00
37 T	Ethyl Acetate	50.000	52.158	-4.3	65	0.00
38 T	Carbon Tetrachloride	50.000	45.280	9.4	59	0.00
39 T	Methylcyclohexane	50.000	45.337	9.3	57	0.00
40 TM	Benzene	50.000	47.989	4.0	61	0.00
41 T	Methacrylonitrile	50.000	50.998	-2.0	62	0.00
42 TM	1,2-Dichloroethane	50.000	46.836	6.3	59	0.00
43 T	Isopropyl Acetate	50.000	48.302	3.4	60	0.00
44 TM	Trichloroethene	50.000	48.926	2.1	62	0.00
45 C	1,2-Dichloropropane	50.000	50.592	-1.2#	64	0.00
46 T	Dibromomethane	50.000	49.937	0.1	63	0.00
47 T	Bromodichloromethane	50.000	51.689	-3.4	63	0.00
48 T	Methyl methacrylate	50.000	47.717	4.6	59	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
 Data File : VX043726.D
 Acq On : 07 Nov 2024 10:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 08 04:02:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	956.867	4.3	59	0.00
50 S	Toluene-d8	50.000	51.716	-3.4	67	0.00
51 T	4-Methyl-2-Pentanone	250.000	256.791	-2.7	62	0.00
52 CM	Toluene	50.000	50.497	-1.0#	63	0.00
53 T	t-1,3-Dichloropropene	50.000	51.097	-2.2	62	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.910	-5.8	64	0.00
55 T	1,1,2-Trichloroethane	50.000	54.381	-8.8	67	0.00
56 T	Ethyl methacrylate	50.000	51.545	-3.1	62	0.00
57 T	1,3-Dichloropropane	50.000	51.876	-3.8	64	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	236.461	5.4	61	0.00
59 T	2-Hexanone	250.000	260.484	-4.2	63	0.00
60 T	Dibromochloromethane	50.000	56.409	-12.8	67	0.00
61 T	1,2-Dibromoethane	50.000	52.726	-5.5	65	0.00
62 S	4-Bromofluorobenzene	50.000	53.995	-8.0	68	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	66	0.00
64 T	Tetrachloroethene	50.000	47.595	4.8	64	0.00
65 PM	Chlorobenzene	50.000	50.064	-0.1	67	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.010	-8.0	68	0.00
67 C	Ethyl Benzene	50.000	47.137	5.7#	62	0.00
68 T	m/p-Xylenes	100.000	101.208	-1.2	65	0.00
69 T	o-Xylene	50.000	51.848	-3.7	67	0.00
70 T	Styrene	50.000	53.257	-6.5	67	0.00
71 P	Bromoform	50.000	54.035	-8.1	66	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	68	0.00
73 T	Isopropylbenzene	50.000	46.842	6.3	63	0.00
74 T	N-amyl acetate	50.000	43.961	12.1	58	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.529	0.9	66	0.00
76 T	1,2,3-Trichloropropane	50.000	49.061	1.9	67	0.00
77 T	Bromobenzene	50.000	51.285	-2.6	69	0.00
78 T	n-propylbenzene	50.000	48.211	3.6	63	0.00
79 T	2-Chlorotoluene	50.000	47.944	4.1	66	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.305	3.4	64	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	51.204	-2.4	65	0.00
82 T	4-Chlorotoluene	50.000	48.785	2.4	66	0.00
83 T	tert-Butylbenzene	50.000	48.296	3.4	65	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.067	-0.1	66	0.00
85 T	sec-Butylbenzene	50.000	48.754	2.5	64	0.00
86 T	p-Isopropyltoluene	50.000	49.115	1.8	64	0.00
87 T	1,3-Dichlorobenzene	50.000	51.900	-3.8	70	0.00
88 T	1,4-Dichlorobenzene	50.000	51.143	-2.3	71	0.00
89 T	n-Butylbenzene	50.000	49.300	1.4	63	0.00
90 T	Hexachloroethane	50.000	49.587	0.8	63	0.00
91 T	1,2-Dichlorobenzene	50.000	52.280	-4.6	69	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.312	3.4	61	0.00
93 T	1,2,4-Trichlorobenzene	50.000	53.772	-7.5	69	0.00
94 T	Hexachlorobutadiene	50.000	48.597	2.8	66	0.00
95 T	Naphthalene	50.000	50.666	-1.3	65	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
Data File : VX043726.D
Acq On : 07 Nov 2024 10:46
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 08 04:02:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	53.557	-7.1	71	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6