



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110724\
Data File : VX043725.D
Acq On : 07 Nov 2024 10:18
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Title : SW846 8260
Last Update : Mon Oct 28 13:41:34 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	4328	PASS
75	95	30	60	46.2	12408	PASS
95	95	100	100	100.0	26880	PASS
96	95	5	9	7.0	1881	PASS
173	174	0.00	2	0.8	163	PASS
174	95	50	100	77.2	20756	PASS
175	174	5	9	7.6	1582	PASS
176	174	95	101	100.3	20826	PASS
177	176	5	9	6.3	1310	PASS