

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX110725\  
 Data File : VX048525.D  
 Acq On : 07 Nov 2025 16:50  
 Operator : JC/MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 ICVVX110725

Quant Time: Nov 08 05:04:12 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X110725W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Nov 08 05:01:56 2025  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By : John Carlone 11/10/2025  
 Supervised By : Semsettin Yesilyurt 11/10/2025

| Compound                     | R.T.           | QIon | Response            | Conc    | Units | Dev(Min) |
|------------------------------|----------------|------|---------------------|---------|-------|----------|
| -----                        |                |      |                     |         |       |          |
| Internal Standards           |                |      |                     |         |       |          |
| 1) Pentafluorobenzene        | 5.544          | 168  | 175429              | 50.000  | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.745          | 114  | 300694              | 50.000  | ug/l  | 0.00     |
| 63) Chlorobenzene-d5         | 10.037         | 117  | 261891              | 50.000  | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.006         | 152  | 133606              | 50.000  | ug/l  | 0.00     |
| System Monitoring Compounds  |                |      |                     |         |       |          |
| 33) 1,2-Dichloroethane-d4    | 5.940          | 65   | 125725              | 51.433  | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 78 - 117 |      | Recovery = 102.860% |         |       |          |
| 35) Dibromofluoromethane     | 5.373          | 113  | 104074              | 45.730  | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = 91.460%  |         |       |          |
| 50) Toluene-d8               | 8.635          | 98   | 360437              | 50.780  | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 92 - 112 |      | Recovery = 101.560% |         |       |          |
| 62) 4-Bromofluorobenzene     | 11.061         | 95   | 134490              | 50.843  | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 83 - 123 |      | Recovery = 101.680% |         |       |          |
| Target Compounds             |                |      |                     |         |       |          |
|                              |                |      |                     |         |       | Qvalue   |
| 2) Dichlorodifluoromethane   | 1.173          | 85   | 79523               | 55.942  | ug/l  | 99       |
| 3) Chloromethane             | 1.301          | 50   | 100103              | 53.094  | ug/l  | 100      |
| 4) Vinyl Chloride            | 1.380          | 62   | 105708              | 49.519  | ug/l  | 96       |
| 5) Bromomethane              | 1.611          | 94   | 74133               | 54.997  | ug/l  | 97       |
| 6) Chloroethane              | 1.691          | 64   | 66991               | 47.919  | ug/l  | 100      |
| 7) Trichlorofluoromethane    | 1.892          | 101  | 163901              | 48.532  | ug/l  | 98       |
| 8) Diethyl Ether             | 2.136          | 74   | 68302               | 50.384  | ug/l  | 99       |
| 9) 1,1,2-Trichlorotrifluo... | 2.337          | 101  | 97010               | 49.618  | ug/l  | 99       |
| 10) Methyl Iodide            | 2.453          | 142  | 148327              | 51.263  | ug/l  | 99       |
| 11) Tert butyl alcohol       | 2.941          | 59   | 119049              | 257.016 | ug/l  | 98       |
| 12) 1,1-Dichloroethene       | 2.325          | 96   | 96593               | 47.299  | ug/l  | 95       |
| 13) Acrolein                 | 2.239          | 56   | 17055               | 254.482 | ug/l  | 98       |
| 14) Allyl chloride           | 2.666          | 41   | 180497              | 47.471  | ug/l  | 99       |
| 15) Acrylonitrile            | 3.056          | 53   | 339497              | 251.074 | ug/l  | 99       |
| 16) Acetone                  | 2.374          | 43   | 289425              | 250.130 | ug/l  | 98       |
| 17) Carbon Disulfide         | 2.514          | 76   | 263284              | 45.815  | ug/l  | 100      |
| 18) Methyl Acetate           | 2.697          | 43   | 151597              | 47.674  | ug/l  | 99       |
| 19) Methyl tert-butyl Ether  | 3.105          | 73   | 366166              | 50.775  | ug/l  | 100      |
| 20) Methylene Chloride       | 2.788          | 84   | 112574              | 47.322  | ug/l  | 93       |
| 21) trans-1,2-Dichloroethene | 3.093          | 96   | 104272              | 48.175  | ug/l  | 97       |
| 22) Diisopropyl ether        | 3.751          | 45   | 365126              | 50.011  | ug/l  | 98       |
| 23) Vinyl Acetate            | 3.715          | 43   | 1623158             | 253.844 | ug/l  | 99       |
| 24) 1,1-Dichloroethane       | 3.605          | 63   | 193671              | 48.344  | ug/l  | 99       |
| 25) 2-Butanone               | 4.538          | 43   | 453513              | 264.945 | ug/l  | 97       |
| 26) 2,2-Dichloropropane      | 4.465          | 77   | 162290              | 47.519  | ug/l  | 100      |
| 27) cis-1,2-Dichloroethene   | 4.483          | 96   | 129253              | 48.854  | ug/l  | 98       |
| 28) Bromochloromethane       | 4.885          | 49   | 97361               | 49.472  | ug/l  | 98       |
| 29) Tetrahydrofuran          | 4.989          | 42   | 307563              | 251.928 | ug/l  | 99       |
| 30) Chloroform               | 5.080          | 83   | 203402              | 49.181  | ug/l  | 99       |
| 31) Cyclohexane              | 5.464          | 56   | 166346              | 50.678  | ug/l  | 98       |
| 32) 1,1,1-Trichloroethane    | 5.373          | 97   | 177144              | 49.079  | ug/l  | 100      |
| 36) 1,1-Dichloropropene      | 5.678          | 75   | 136129              | 44.670  | ug/l  | 98       |
| 37) Ethyl Acetate            | 4.696          | 43   | 206366              | 48.107  | ug/l  | 98       |
| 38) Carbon Tetrachloride     | 5.666          | 117  | 153372              | 44.212  | ug/l  | 97       |
| 39) Methylcyclohexane        | 7.367          | 83   | 168487              | 49.240  | ug/l  | 98       |
| 40) Benzene                  | 6.025          | 78   | 428574              | 45.369  | ug/l  | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX110725\  
 Data File : VX048525.D  
 Acq On : 07 Nov 2025 16:50  
 Operator : JC/MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 ICVVX110725

Quant Time: Nov 08 05:04:12 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X110725W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Nov 08 05:01:56 2025  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By : John Carlone 11/10/2025  
 Supervised By : Semsettin Yesilyurt 11/10/2025

| Compound                      | R.T.   | QIon | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|----------|--------|----------|
| 41) Methacrylonitrile         | 4.904  | 41   | 96023    | 50.900   | ug/l   | 95       |
| 42) 1,2-Dichloroethane        | 6.074  | 62   | 149489   | 46.187   | ug/l   | 99       |
| 43) Isopropyl Acetate         | 6.324  | 43   | 292708   | 48.542   | ug/l   | 99       |
| 44) Trichloroethene           | 7.111  | 130  | 108399   | 48.007   | ug/l   | 95       |
| 45) 1,2-Dichloropropane       | 7.415  | 63   | 109605   | 50.462   | ug/l   | 97       |
| 46) Dibromomethane            | 7.568  | 93   | 81296    | 51.317   | ug/l   | 97       |
| 47) Bromodichloromethane      | 7.806  | 83   | 164733   | 52.124   | ug/l   | 99       |
| 48) Methyl methacrylate       | 7.677  | 41   | 141804   | 52.901   | ug/l   | 99       |
| 49) 1,4-Dioxane               | 7.647  | 88   | 43542    | 1066.027 | ug/l   | 97       |
| 51) 4-Methyl-2-Pentanone      | 8.555  | 43   | 943143   | 274.770  | ug/l   | 99       |
| 52) Toluene                   | 8.702  | 92   | 263658   | 53.014   | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 8.964  | 75   | 172732   | 53.042   | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene   | 8.354  | 75   | 183768   | 54.498   | ug/l   | 99       |
| 55) 1,1,2-Trichloroethane     | 9.135  | 97   | 109093   | 54.903   | ug/l   | 98       |
| 56) Ethyl methacrylate        | 9.098  | 69   | 186912   | 56.539   | ug/l   | 98       |
| 57) 1,3-Dichloropropane       | 9.293  | 76   | 182103   | 52.652   | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 8.226  | 63   | 533039   | 299.907  | ug/l   | 100      |
| 59) 2-Hexanone                | 9.415  | 43   | 708416   | 277.453  | ug/l   | 99       |
| 60) Dibromochloromethane      | 9.506  | 129  | 130424   | 53.460   | ug/l   | 100      |
| 61) 1,2-Dibromoethane         | 9.592  | 107  | 115352   | 53.598   | ug/l   | 99       |
| 64) Tetrachloroethene         | 9.256  | 164  | 91541    | 48.124   | ug/l   | 96       |
| 65) Chlorobenzene             | 10.061 | 112  | 292507   | 47.057   | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 10.147 | 131  | 103725   | 49.801   | ug/l   | 99       |
| 67) Ethyl Benzene             | 10.177 | 91   | 505794   | 48.456   | ug/l   | 97       |
| 68) m/p-Xylenes               | 10.287 | 106  | 391090   | 99.482   | ug/l   | 98       |
| 69) o-Xylene                  | 10.622 | 106  | 190122   | 49.430   | ug/l   | 98       |
| 70) Styrene                   | 10.640 | 104  | 318495   | 49.362   | ug/l   | 99       |
| 71) Bromoform                 | 10.781 | 173  | 92512    | 48.145   | ug/l # | 100      |
| 73) Isopropylbenzene          | 10.945 | 105  | 479112   | 49.403   | ug/l   | 99       |
| 74) N-amyl acetate            | 10.823 | 43   | 244571   | 50.225   | ug/l   | 98       |
| 75) 1,1,2,2-Tetrachloroethane | 11.195 | 83   | 167085   | 48.855   | ug/l   | 99       |
| 76) 1,2,3-Trichloropropane    | 11.220 | 75   | 133852m  | 41.585   | ug/l   |          |
| 77) Bromobenzene              | 11.183 | 156  | 120400   | 48.079   | ug/l   | 98       |
| 78) n-propylbenzene           | 11.287 | 91   | 560609   | 49.162   | ug/l   | 100      |
| 79) 2-Chlorotoluene           | 11.348 | 91   | 336034   | 48.767   | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 11.433 | 105  | 396355   | 50.134   | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 11.000 | 75   | 57267    | 50.890   | ug/l   | 96       |
| 82) 4-Chlorotoluene           | 11.433 | 91   | 396344   | 49.035   | ug/l   | 100      |
| 83) tert-Butylbenzene         | 11.695 | 119  | 406392   | 49.698   | ug/l   | 98       |
| 84) 1,2,4-Trimethylbenzene    | 11.732 | 105  | 399935   | 50.137   | ug/l   | 99       |
| 85) sec-Butylbenzene          | 11.872 | 105  | 498225   | 49.297   | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 11.988 | 119  | 419888   | 49.417   | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 11.951 | 146  | 223034   | 47.519   | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 12.024 | 146  | 225606   | 46.536   | ug/l   | 99       |
| 89) n-Butylbenzene            | 12.311 | 91   | 390447   | 48.984   | ug/l   | 100      |
| 90) Hexachloroethane          | 12.518 | 117  | 78393    | 48.425   | ug/l   | 100      |
| 91) 1,2-Dichlorobenzene       | 12.317 | 146  | 216201   | 47.984   | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.920 | 75   | 40188    | 50.566   | ug/l   | 99       |
| 93) 1,2,4-Trichlorobenzene    | 13.567 | 180  | 148227   | 49.507   | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 13.707 | 225  | 58865    | 47.031   | ug/l   | 99       |
| 95) Naphthalene               | 13.756 | 128  | 506116   | 52.576   | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.939 | 180  | 140560   | 50.471   | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX110725\  
Data File : VX048525.D  
Acq On : 07 Nov 2025 16:50  
Operator : JC/MD  
Sample : VSTDICV050  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX110725

Manual Integrations  
APPROVED

Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 05:04:12 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X110725W.M  
Quant Title : SW846 8260  
QLast Update : Sat Nov 08 05:01:56 2025  
Response via : Initial Calibration


| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

## Manual Integrations

Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025

[illegible]







#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.544 min Scan# 71  
Delta R.T. 0.006 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX110725

Tgt Ion:168 Resp: 175429  
Ion Ratio Lower Upper  
168 100  
99 59.3 45.2 67.8




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#2  
Dichlorodifluoromethane  
Concen: 55.942 ug/l  
RT: 1.173 min Scan# 14  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50


Tgt Ion: 85 Resp: 79523  
Ion Ratio Lower Upper  
85 100  
87 31.9 16.2 48.6





#3  
Chloromethane  
Concen: 53.094 ug/l  
RT: 1.301 min Scan# 35  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVX110725



Tgt Ion: 50 Resp: 10010  
Ion Ratio Lower Upper  
50 100  
52 32.3 26.0 39.0




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#4  
Vinyl Chloride  
Concen: 49.519 ug/l  
RT: 1.380 min Scan# 48  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Tgt Ion: 62 Resp: 105708  
Ion Ratio Lower Upper  
62 100  
64 30.8 26.2 39.4









#5  
Bromomethane  
Concen: 54.997 ug/l  
RT: 1.611 min Scan# 86  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Instrument :  
MSVOA\_X  
Client Sample Id :  
ICVX110725

Tgt Ion: 94 Resp: 7413  
Ion Ratio Lower Upper  
94 100  
96 93.0 72.3 108.5




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#6  
Chloroethane  
Concen: 47.919 ug/l  
RT: 1.691 min Scan# 99  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Tgt Ion: 64 Resp: 66991  
Ion Ratio Lower Upper  
64 100  
66 32.2 25.9 38.9









#7  
 Trichlorofluoromethane  
 Concen: 48.532 ug/l  
 RT: 1.892 min Scan# 11  
 Delta R.T. -0.000 min  
 Lab File: VX048525.D  
 Acq: 07 Nov 2025 16:50

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 ICVVX110725

Tgt Ion:101 Resp: 16390  
 Ion Ratio Lower Upper  
 101 100  
 103 63.4 51.7 77.5




Manual Integrations  
 APPROVED


Reviewed By :John Carlone 11/10/2025  
 Supervised By :Semsettin Yesilyurt 11/10/2025



#8  
 Diethyl Ether  
 Concen: 50.384 ug/l  
 RT: 2.136 min Scan# 172  
 Delta R.T. -0.000 min  
 Lab File: VX048525.D  
 Acq: 07 Nov 2025 16:50

Tgt Ion: 74 Resp: 68302  
 Ion Ratio Lower Upper  
 74 100  
 45 92.8 45.9 137.6









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 49.618 ug/l  
RT: 2.337 min Scan# 204  
Delta R.T. 0.006 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVX110725

Tgt Ion:101 Resp: 97010  
Ion Ratio Lower Upper  
101 100  
85 42.4 34.3 51.5  
151 78.2 63.1 94.7




Manual Integrations  
APPROVED

Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#10  
Methyl Iodide  
Concen: 51.263 ug/l  
RT: 2.453 min Scan# 224  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Tgt Ion:142 Resp: 148327  
Ion Ratio Lower Upper  
142 100  
127 44.2 35.8 53.6  
141 14.1 11.4 17.0





#11

Tert butyl alcohol

Concen: 257.016 ug/l

RT: 2.941 min Scan# 304

Delta R.T. -0.000 min

Lab File: VX048525.D

Acq: 07 Nov 2025 16:50

Instrument :

MSVOA\_X

ClientSampleId :

ICVVX110725

Tgt Ion: 59 Resp: 119049

Ion Ratio Lower Upper

59 100





57 10.1 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 11/10/2025

Supervised By :Semsettin Yesilyurt 11/10/2025



#12

1,1-Dichloroethene

Concen: 47.299 ug/l

RT: 2.325 min Scan# 203

Delta R.T. -0.000 min

Lab File: VX048525.D

Acq: 07 Nov 2025 16:50




Tgt Ion: 96 Resp: 96593


Ion Ratio Lower Upper

96 100

61 153.7 129.8 194.6

98 63.4 52.3 78.5









#13  
Acrolein  
Concen: 254.482 ug/l  
RT: 2.239 min Scan# 189  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVX110725

Tgt Ion: 56 Resp: 1705  
Ion Ratio Lower Upper  
56 100  
55 70.4 54.9 82.3




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#14  
Allyl chloride  
Concen: 47.471 ug/l  
RT: 2.666 min Scan# 259  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50


Tgt Ion: 41 Resp: 180497  
Ion Ratio Lower Upper  
41 100  
39 62.0 48.6 73.0  
76 34.7 27.7 41.5





#15  
Acrylonitrile  
Concen: 251.074 ug/l  
RT: 3.056 min Scan# 31  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVX110725



Tgt Ion: 53 Resp: 33949  
Ion Ratio Lower Upper  
53 100  
52 82.0 65.3 97.9  
51 35.3 27.8 41.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#16  
Acetone  
Concen: 250.130 ug/l  
RT: 2.374 min Scan# 211  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50


Tgt Ion: 43 Resp: 289425  
Ion Ratio Lower Upper  
43 100  
58 31.3 25.8 38.6





#17  
Carbon Disulfide  
Concen: 45.815 ug/l  
RT: 2.514 min Scan# 21  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX110725



Tgt Ion: 76 Resp: 263284  
Ion Ratio Lower Upper  
76 100  
78 9.0 7.1 10.7




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#18  
Methyl Acetate  
Concen: 47.674 ug/l  
RT: 2.697 min Scan# 264  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50


Tgt Ion: 43 Resp: 151597  
Ion Ratio Lower Upper  
43 100  
74 24.6 19.1 28.7





#19  
Methyl tert-butyl Ether  
Concen: 50.775 ug/l  
RT: 3.105 min Scan# 331  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX110725



Tgt Ion: 73 Resp: 366160  
Ion Ratio Lower Upper  
73 100  
57 21.6 17.3 25.9




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#20  
Methylene Chloride  
Concen: 47.322 ug/l  
RT: 2.788 min Scan# 279  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50


Tgt Ion: 84 Resp: 112574  
Ion Ratio Lower Upper  
84 100  
49 123.0 91.2 136.8  
51 35.8 28.3 42.5  
86 67.2 49.4 74.2





#21  
trans-1,2-Dichloroethene  
Concen: 48.175 ug/l  
RT: 3.093 min Scan# 31  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVX110725



Tgt Ion: 96 Resp: 104272  
Ion Ratio Lower Upper  
96 100  
61 129.6 106.7 160.1  
98 63.7 52.7 79.1




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#22  
Diisopropyl ether  
Concen: 50.011 ug/l  
RT: 3.751 min Scan# 437  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50


Tgt Ion: 45 Resp: 365126  
Ion Ratio Lower Upper  
45 100  
43 123.0 96.4 144.6  
87 28.1 21.8 32.6  
59 11.5 9.6 14.4





#23  
 Vinyl Acetate  
 Concen: 253.844 ug/l  
 RT: 3.715 min Scan# 413  
 Delta R.T. -0.000 min  
 Lab File: VX048525.D  
 Acq: 07 Nov 2025 16:50




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 ICVVX110725




Tgt Ion: 43 Resp: 1623158  
 Ion Ratio Lower Upper  
 43 100  
 86 11.1 9.0 13.6

Manual Integrations  
 APPROVED



Reviewed By :John Carlone 11/10/2025  
 Supervised By :Semsettin Yesilyurt 11/10/2025




#24  
 1,1-Dichloroethane  
 Concen: 48.344 ug/l  
 RT: 3.605 min Scan# 413  
 Delta R.T. -0.000 min  
 Lab File: VX048525.D  
 Acq: 07 Nov 2025 16:50



Tgt Ion: 63 Resp: 193671  
 Ion Ratio Lower Upper  
 63 100  
 98 6.9 3.4 10.2  
 100 4.6 2.5 7.4









#25  
2-Butanone  
Concen: 264.945 ug/l  
RT: 4.538 min Scan# 50  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX110725

Tgt Ion: 43 Resp: 45351  
Ion Ratio Lower Upper  
43 100  
72 25.2 21.5 32.3




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#26  
2,2-Dichloropropane  
Concen: 47.519 ug/l  
RT: 4.465 min Scan# 554  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50


Tgt Ion: 77 Resp: 162290  
Ion Ratio Lower Upper  
77 100  
97 23.4 11.7 35.1





#27  
 cis-1,2-Dichloroethene  
 Concen: 48.854 ug/l  
 RT: 4.483 min Scan# 557  
 Delta R.T. -0.000 min  
 Lab File: VX048525.D  
 Acq: 07 Nov 2025 16:50




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 ICVVX110725



Tgt Ion: 96 Resp: 12925  
 Ion Ratio Lower Upper  
 96 100  
 61 142.8 0.0 279.6  
 98 64.8 0.0 128.2




Manual Integrations  
 APPROVED


Reviewed By :John Carlone 11/10/2025  
 Supervised By :Semsettin Yesilyurt 11/10/2025



#28  
 Bromochloromethane  
 Concen: 49.472 ug/l  
 RT: 4.885 min Scan# 623  
 Delta R.T. -0.000 min  
 Lab File: VX048525.D  
 Acq: 07 Nov 2025 16:50

Tgt Ion: 49 Resp: 97361  
 Ion Ratio Lower Upper  
 49 100  
 129 2.1 0.0 4.2  
 130 80.8 65.9 98.9





#29

Tetrahydrofuran

Concen: 251.928 ug/l

RT: 4.989 min Scan# 640

Delta R.T. -0.000 min

Lab File: VX048525.D

Acq: 07 Nov 2025 16:50

Instrument :

MSVOA\_X

ClientSampleId :

ICVX110725

Tgt Ion: 42 Resp: 30756

Ion Ratio Lower Upper

42 100

72 44.6 36.3 54.5





71 41.6 34.1 51.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 11/10/2025

Supervised By :Semsettin Yesilyurt 11/10/2025



#30

Chloroform

Concen: 49.181 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX048525.D




Acq: 07 Nov 2025 16:50


Tgt Ion: 83 Resp: 203402

Ion Ratio Lower Upper

83 100


85 64.1 51.8 77.6





#31  
Cyclohexane  
Concen: 50.678 ug/l  
RT: 5.464 min Scan# 718  
Delta R.T. 0.006 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVX110725



Tgt Ion: 56 Resp: 166340  
Ion Ratio Lower Upper  
56 100  
69 30.9 23.4 35.2  
84 84.2 66.3 99.5




Manual Integrations  
APPROVED

Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#32  
1,1,1-Trichloroethane  
Concen: 49.079 ug/l  
RT: 5.373 min Scan# 703  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Tgt Ion: 97 Resp: 177144  
Ion Ratio Lower Upper  
97 100  
99 63.4 50.8 76.2  
61 41.8 33.4 50.2





#33

1,2-Dichloroethane-d4

Concen: 51.433 ug/l

RT: 5.940 min Scan# 797

Delta R.T. -0.006 min

Lab File: VX048525.D

Acq: 07 Nov 2025 16:50

Instrument :

MSVOA\_X

ClientSampleId :

ICVVX110725

Tgt Ion: 65 Resp: 125725

Ion Ratio Lower Upper

65 100





67 53.7 0.0 106.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 11/10/2025

Supervised By :Semsettin Yesilyurt 11/10/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.745 min Scan# 928

Delta R.T. -0.000 min

Lab File: VX048525.D

Acq: 07 Nov 2025 16:50




Tgt Ion:114 Resp: 300694


Ion Ratio Lower Upper

114 100

63 21.4 0.0 41.2


88 16.6 0.0 32.2





#35  
Dibromofluoromethane  
Concen: 45.730 ug/l  
RT: 5.373 min Scan# 703  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVX110725



Tgt Ion: 113 Resp: 104074  
Ion Ratio Lower Upper  
113 100  
111 101.7 81.9 122.9  
192 19.6 15.8 23.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#36  
1,1-Dichloropropene  
Concen: 44.670 ug/l  
RT: 5.678 min Scan# 753  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Tgt Ion: 75 Resp: 136129  
Ion Ratio Lower Upper  
75 100  
110 35.4 18.4 55.4  
77 30.3 24.6 37.0





#37

Ethyl Acetate

Concen: 48.107 ug/l

RT: 4.696 min Scan# 592

Delta R.T. -0.000 min

Lab File: VX048525.D

Acq: 07 Nov 2025 16:50

Instrument :

MSVOA\_X

ClientSampleId :

ICVX110725

Tgt Ion: 43 Resp: 206360

Ion Ratio Lower Upper

43 100

61 11.0 9.7 14.5





70 9.3 7.7 11.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 11/10/2025

Supervised By :Semsettin Yesilyurt 11/10/2025



#38

Carbon Tetrachloride

Concen: 44.212 ug/l

RT: 5.666 min Scan# 751

Delta R.T. -0.000 min

Lab File: VX048525.D

Acq: 07 Nov 2025 16:50




Tgt Ion:117 Resp: 153372


Ion Ratio Lower Upper

117 100

119 96.2 74.5 111.7

121 30.8 24.8 37.2





#39

Methylcyclohexane

Concen: 49.240 ug/l

RT: 7.367 min Scan# 1030

Delta R.T. -0.000 min

Lab File: VX048525.D

Acq: 07 Nov 2025 16:50

Instrument :

MSVOA\_X

ClientSampleId :

ICVX110725

Tgt Ion: 83 Resp: 16848

Ion Ratio Lower Upper

83 100

55 77.6 63.8 95.6





98 49.6 40.3 60.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 11/10/2025

Supervised By :Semsettin Yesilyurt 11/10/2025



#40

Benzene

Concen: 45.369 ug/l

RT: 6.025 min Scan# 810

Delta R.T. -0.000 min

Lab File: VX048525.D




Acq: 07 Nov 2025 16:50


Tgt Ion: 78 Resp: 428574

Ion Ratio Lower Upper

78 100

77 23.6 18.3 27.5






#41  
Methacrylonitrile  
Concen: 50.900 ug/l  
RT: 4.904 min Scan# 61  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Instrument :

MSVOA\_X

ClientSampleId :

ICVVX110725






Tgt Ion: 41 Resp: 9602  
Ion Ratio Lower Upper  
41 100  
39 52.2 41.1 61.7  
67 75.3 64.2 96.2  
52 26.9 25.6 38.4

Manual Integrations

APPROVED




Reviewed By :John Carlone 11/10/2025


Supervised By :Semsettin Yesilyurt 11/10/2025



#42  
1,2-Dichloroethane  
Concen: 46.187 ug/l  
RT: 6.074 min Scan# 818  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Tgt Ion: 62 Resp: 149489  
Ion Ratio Lower Upper  
62 100  
98 9.5 0.0 19.6






#43  
Isopropyl Acetate  
Concen: 48.542 ug/l  
RT: 6.324 min Scan# 859  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Instrument :

MSVOA\_X

Client Sample Id :

ICVX110725






Tgt Ion: 43 Resp: 292708  
Ion Ratio Lower Upper  
43 100  
61 20.1 16.2 24.4  
87 13.0 10.9 16.3

Manual Integrations

APPROVED




Reviewed By :John Carlone 11/10/2025


Supervised By :Semsettin Yesilyurt 11/10/2025



#44  
Trichloroethene  
Concen: 48.007 ug/l  
RT: 7.111 min Scan# 988  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50


Tgt Ion:130 Resp: 108399  
Ion Ratio Lower Upper  
130 100  
95 100.1 0.0 190.2





#45  
1,2-Dichloropropane  
Concen: 50.462 ug/l  
RT: 7.415 min Scan# 1038  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVX110725



Tgt Ion: 63 Resp: 10960  
Ion Ratio Lower Upper  
63 100  
65 29.5 24.9 37.3




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#46  
Dibromomethane  
Concen: 51.317 ug/l  
RT: 7.568 min Scan# 1063  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50


Tgt Ion: 93 Resp: 81296  
Ion Ratio Lower Upper  
93 100  
95 82.3 67.6 101.4  
174 100.5 83.0 124.6





#47  
Bromodichloromethane  
Concen: 52.124 ug/l  
RT: 7.806 min Scan# 1102  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX110725



Tgt Ion: 83 Resp: 164733  
Ion Ratio Lower Upper  
83 100  
85 63.3 51.4 77.0  
127 7.6 7.0 10.4




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#48  
Methyl methacrylate  
Concen: 52.901 ug/l  
RT: 7.677 min Scan# 1081  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Tgt Ion: 41 Resp: 141804  
Ion Ratio Lower Upper  
41 100  
69 83.8 67.9 101.9  
39 53.8 42.2 63.2





#49

1,4-Dioxane

Concen: 1066.027 ug/l

RT: 7.647 min Scan# 1076

Delta R.T. -0.000 min

Lab File: VX048525.D

Acq: 07 Nov 2025 16:50

Instrument :

MSVOA\_X

ClientSampleId :

ICVX110725

Tgt Ion: 88 Resp: 4354





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 88  | 100   |       |       |
| 43  | 35.8  | 25.1  | 37.7  |
| 58  | 70.5  | 56.7  | 85.1  |

Manual Integrations

APPROVED

Reviewed By :John Carlone 11/10/2025

Supervised By :Semsettin Yesilyurt 11/10/2025



#50

Toluene-d8

Concen: 50.780 ug/l

RT: 8.635 min Scan# 1238




Delta R.T. -0.000 min


Lab File: VX048525.D

Acq: 07 Nov 2025 16:50

Tgt Ion: 98 Resp: 360437


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 98  | 100   |       |       |
| 100 | 67.5  | 53.6  | 80.4  |





#51  
4-Methyl-2-Pentanone  
Concen: 274.770 ug/l  
RT: 8.555 min Scan# 1225  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX110725



Tgt Ion: 43 Resp: 94314  
Ion Ratio Lower Upper  
43 100  
58 39.9 32.3 48.5




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#52  
Toluene  
Concen: 53.014 ug/l  
RT: 8.702 min Scan# 1249  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Tgt Ion: 92 Resp: 263658  
Ion Ratio Lower Upper  
92 100  
91 172.6 137.0 205.6





#53

t-1,3-Dichloropropene

Concen: 53.042 ug/l

RT: 8.964 min Scan# 1192

Delta R.T. -0.000 min

Lab File: VX048525.D

Acq: 07 Nov 2025 16:50

Instrument :

MSVOA\_X

ClientSampleId :

ICVX110725

Tgt Ion: 75 Resp: 17273

Ion Ratio Lower Upper

75 100





77 31.8 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 11/10/2025

Supervised By :Semsettin Yesilyurt 11/10/2025



#54

cis-1,3-Dichloropropene

Concen: 54.498 ug/l

RT: 8.354 min Scan# 1192

Delta R.T. -0.000 min

Lab File: VX048525.D

Acq: 07 Nov 2025 16:50




Tgt Ion: 75 Resp: 183768


Ion Ratio Lower Upper

75 100

77 31.8 24.7 37.1


39 44.9 36.1 54.1





#55  
1,1,2-Trichloroethane  
Concen: 54.903 ug/l  
RT: 9.135 min Scan# 1320  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument : MSVOA\_X  
ClientSampleId : ICVVX110725



Tgt Ion: 97 Resp: 10909  
Ion Ratio Lower Upper  
97 100  
83 83.9 69.2 103.8  
85 54.8 42.8 64.2  
99 61.9 49.7 74.5




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#56  
Ethyl methacrylate  
Concen: 56.539 ug/l  
RT: 9.098 min Scan# 1314  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Tgt Ion: 69 Resp: 186912  
Ion Ratio Lower Upper  
69 100  
41 67.6 53.3 79.9  
39 34.3 26.1 39.1





#57

1,3-Dichloropropane

Concen: 52.652 ug/l

RT: 9.293 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX048525.D

Acq: 07 Nov 2025 16:50

Instrument :

MSVOA\_X

ClientSampleId :

ICVX110725

Tgt Ion: 76 Resp: 18210

Ion Ratio Lower Upper

76 100





78 32.5 26.2 39.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 11/10/2025

Supervised By :Semsettin Yesilyurt 11/10/2025



#58

2-Chloroethyl Vinyl ether

Concen: 299.907 ug/l

RT: 8.226 min Scan# 1171

Delta R.T. -0.000 min

Lab File: VX048525.D




Acq: 07 Nov 2025 16:50


Tgt Ion: 63 Resp: 533039

Ion Ratio Lower Upper

63 100


106 28.5 23.0 34.4





#59  
2-Hexanone  
Concen: 277.453 ug/l  
RT: 9.415 min Scan# 1366  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVX110725



Tgt Ion: 43 Resp: 708410  
Ion Ratio Lower Upper  
43 100  
58 55.0 28.1 84.2




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#60  
Dibromochloromethane  
Concen: 53.460 ug/l  
RT: 9.506 min Scan# 1381  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Tgt Ion:129 Resp: 130424  
Ion Ratio Lower Upper  
129 100  
127 75.8 38.0 114.0









#61  
1,2-Dibromoethane  
Concen: 53.598 ug/l  
RT: 9.592 min Scan# 1395  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVX110725

Tgt Ion:107 Resp: 115352  
Ion Ratio Lower Upper  
107 100  
109 93.4 75.3 112.9




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#62  
4-Bromofluorobenzene  
Concen: 50.843 ug/l  
RT: 11.061 min Scan# 1636  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50


Tgt Ion: 95 Resp: 134490  
Ion Ratio Lower Upper  
95 100  
174 77.9 0.0 161.8  
176 76.0 0.0 153.6





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.037 min Scan# 1468  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument : MSVOA\_X  
Client Sample Id : ICVVX110725



Tgt Ion:117 Resp: 26189  
Ion Ratio Lower Upper  
117 100  
82 56.8 44.4 66.6  
119 31.4 25.1 37.7




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#64  
Tetrachloroethene  
Concen: 48.124 ug/l  
RT: 9.256 min Scan# 1340  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50


Tgt Ion:164 Resp: 91541  
Ion Ratio Lower Upper  
164 100  
166 121.3 100.6 150.8  
129 91.9 77.6 116.4  
131 89.8 73.0 109.6





#65  
Chlorobenzene  
Concen: 47.057 ug/l  
RT: 10.061 min Scan# 1472  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVX110725




Tgt Ion:112 Resp: 29250  
Ion Ratio Lower Upper  
112 100  
114 31.5 25.5 38.3

Manual Integrations  
APPROVED



Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025




#66  
1,1,1,2-Tetrachloroethane  
Concen: 49.801 ug/l  
RT: 10.147 min Scan# 1486  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Tgt Ion:131 Resp: 103725  
Ion Ratio Lower Upper  
131 100  
133 94.7 47.9 143.7  
119 66.2 32.8 98.3





#67  
Ethyl Benzene  
Concen: 48.456 ug/l  
RT: 10.177 min Scan# 1491  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument : MSVOA\_X  
ClientSampleId : ICVVX110725




Tgt Ion: 91 Resp: 505794  
Ion Ratio Lower Upper  
91 100  
106 31.7 24.1 36.1

Manual Integrations  
APPROVED



Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025




#68  
m/p-Xylenes  
Concen: 99.482 ug/l  
RT: 10.287 min Scan# 1509  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50



Tgt Ion:106 Resp: 391090  
Ion Ratio Lower Upper  
106 100  
91 202.1 159.6 239.4









#69  
o-Xylene  
Concen: 49.430 ug/l  
RT: 10.622 min Scan# 1564  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVX110725

Tgt Ion:106 Resp: 19012  
Ion Ratio Lower Upper  
106 100  
91 209.9 106.4 319.2




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#70  
Styrene  
Concen: 49.362 ug/l  
RT: 10.640 min Scan# 1567  
Delta R.T. 0.006 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50


Tgt Ion:104 Resp: 318495  
Ion Ratio Lower Upper  
104 100  
78 50.1 39.4 59.2  
103 54.1 43.1 64.7





#71  
Bromoform  
Concen: 48.145 ug/l  
RT: 10.781 min Scan# 1590  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVX110725




Tgt Ion:173 Resp: 92512  
Ion Ratio Lower Upper  
173 100  
175 48.6 24.4 73.4  
254 0.0 0.1 0.1

Manual Integrations  
APPROVED



Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.006 min Scan# 1791  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50



Tgt Ion:152 Resp: 133606  
Ion Ratio Lower Upper  
152 100  
115 83.0 39.2 117.6  
150 175.0 0.0 347.0





#73  
Isopropylbenzene  
Concen: 49.403 ug/l  
RT: 10.945 min Scan# 1617  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Instrument :

MSVOA\_X

ClientSampleId :





ICVX110725

Tgt Ion:105 Resp: 47911  
Ion Ratio Lower Upper  
105 100  
120 26.6 13.6 40.7

Manual Integrations




APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#74  
N-ethyl acetate  
Concen: 50.225 ug/l  
RT: 10.823 min Scan# 1597  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Tgt Ion: 43 Resp: 244571  
Ion Ratio Lower Upper  
43 100  
70 42.3 35.1 52.7  
55 23.6 19.3 28.9  
61 24.4 19.9 29.9









#75  
1,1,2,2-Tetrachloroethane  
Concen: 48.855 ug/l  
RT: 11.195 min Scan# 1658  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX110725

Tgt Ion: 83 Resp: 167085  
Ion Ratio Lower Upper  
83 100  
131 10.1 5.3 15.8  
85 64.0 32.3 96.8




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#76  
1,2,3-Trichloropropane  
Concen: 41.585 ug/l m  
RT: 11.220 min Scan# 1662  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Tgt Ion: 75 Resp: 133852  
Ion Ratio Lower Upper  
75 100  
77 49.7 24.3 72.8





#77

Bromobenzene

Concen: 48.079 ug/l

RT: 11.183 min Scan# 1655

Delta R.T. 0.006 min

Lab File: VX048525.D

Acq: 07 Nov 2025 16:50

Instrument :

MSVOA\_X

ClientSampleId :

ICVX110725

Tgt Ion:156 Resp: 120400

Ion Ratio Lower Upper

156 100

77 160.3 78.1 234.4





158 98.2 48.5 145.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 11/10/2025

Supervised By :Semsettin Yesilyurt 11/10/2025



#78

n-propylbenzene

Concen: 49.162 ug/l

RT: 11.287 min Scan# 1673

Delta R.T. -0.000 min

Lab File: VX048525.D




Acq: 07 Nov 2025 16:50


Tgt Ion: 91 Resp: 560609

Ion Ratio Lower Upper

91 100


120 23.2 11.6 34.8





#79  
2-Chlorotoluene  
Concen: 48.767 ug/l  
RT: 11.348 min Scan# 1683  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX110725




Tgt Ion: 91 Resp: 336034  
Ion Ratio Lower Upper  
91 100  
126 33.5 17.1 51.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025




#80  
1,3,5-Trimethylbenzene  
Concen: 50.134 ug/l  
RT: 11.433 min Scan# 1697  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50



Tgt Ion:105 Resp: 396355  
Ion Ratio Lower Upper  
105 100  
120 49.3 25.1 75.2









#81  
trans-1,4-Dichloro-2-butene  
Concen: 50.890 ug/l  
RT: 11.000 min Scan# 1626  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVX110725

Tgt Ion: 75 Resp: 5726  
Ion Ratio Lower Upper  
75 100  
53 95.6 72.4 108.6  
89 53.5 43.6 65.4




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#82  
4-Chlorotoluene  
Concen: 49.035 ug/l  
RT: 11.433 min Scan# 1697  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50


Tgt Ion: 91 Resp: 396344  
Ion Ratio Lower Upper  
91 100  
126 29.8 14.9 44.7





#83  
 tert-Butylbenzene  
 Concen: 49.698 ug/l  
 RT: 11.695 min Scan# 1740  
 Delta R.T. -0.000 min  
 Lab File: VX048525.D  
 Acq: 07 Nov 2025 16:50




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 ICVVX110725




Tgt Ion:119 Resp: 406392  
 Ion Ratio Lower Upper  
 119 100  
 91 56.1 27.1 81.3  
 134 23.0 11.4 34.1

Manual Integrations  
 APPROVED



Reviewed By :John Carlone 11/10/2025  
 Supervised By :Semsettin Yesilyurt 11/10/2025




#84  
 1,2,4-Trimethylbenzene  
 Concen: 50.137 ug/l  
 RT: 11.732 min Scan# 1746  
 Delta R.T. -0.000 min  
 Lab File: VX048525.D  
 Acq: 07 Nov 2025 16:50




Tgt Ion:105 Resp: 399935  
 Ion Ratio Lower Upper  
 105 100  
 120 45.2 22.4 67.0





#85  
sec-Butylbenzene  
Concen: 49.297 ug/l  
RT: 11.872 min Scan# 1769  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX110725



Tgt Ion:105 Resp: 498225  
Ion Ratio Lower Upper  
105 100  
134 19.9 10.2 30.4




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#86  
p-Isopropyltoluene  
Concen: 49.417 ug/l  
RT: 11.988 min Scan# 1788  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50


Tgt Ion:119 Resp: 419888  
Ion Ratio Lower Upper  
119 100  
134 26.6 13.2 39.5  
91 23.3 11.3 33.9





#87  
1,3-Dichlorobenzene  
Concen: 47.519 ug/l  
RT: 11.951 min Scan# 1782  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX110725




Tgt Ion:146 Resp: 223034  
Ion Ratio Lower Upper  
146 100  
111 40.5 19.8 59.4  
148 64.2 31.6 95.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025




#88  
1,4-Dichlorobenzene  
Concen: 46.536 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50



Tgt Ion:146 Resp: 225606  
Ion Ratio Lower Upper  
146 100  
111 39.5 19.2 57.6  
148 63.9 31.9 95.9









#89  
n-Butylbenzene  
Concen: 48.984 ug/l  
RT: 12.311 min Scan# 1841  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Instrument : MSVOA\_X  
Client Sample Id : ICVVX110725

Tgt Ion: 91 Resp: 39044  
Ion Ratio Lower Upper  
91 100  
92 54.4 27.2 81.5  
134 25.7 13.0 39.0




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#90  
Hexachloroethane  
Concen: 48.425 ug/l  
RT: 12.518 min Scan# 1875  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Tgt Ion:117 Resp: 78393  
Ion Ratio Lower Upper  
117 100  
201 72.0 35.9 107.8









#91  
1,2-Dichlorobenzene  
Concen: 47.984 ug/l  
RT: 12.317 min Scan# 1842  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Instrument : MSVOA\_X  
ClientSampleId : ICVVX110725

Tgt Ion:146 Resp: 21620  
Ion Ratio Lower Upper  
146 100  
111 42.1 20.6 61.9  
148 64.8 31.4 94.2




Manual Integrations  
APPROVED


Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#92  
1,2-Dibromo-3-Chloropropane  
Concen: 50.566 ug/l  
RT: 12.920 min Scan# 1941  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Tgt Ion: 75 Resp: 40188  
Ion Ratio Lower Upper  
75 100  
155 87.0 43.5 130.7  
157 108.3 54.8 164.4





#93

1,2,4-Trichlorobenzene

Concen: 49.507 ug/l

RT: 13.567 min Scan# 2047

Delta R.T. -0.000 min

Lab File: VX048525.D

Acq: 07 Nov 2025 16:50

Instrument :

MSVOA\_X

ClientSampleId :

ICVX110725

Tgt Ion:180 Resp: 14822

Ion Ratio Lower Upper

180 100

182 95.3 47.5 142.6





145 30.5 15.9 47.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 11/10/2025

Supervised By :Semsettin Yesilyurt 11/10/2025



#94

Hexachlorobutadiene

Concen: 47.031 ug/l

RT: 13.707 min Scan# 2070

Delta R.T. -0.000 min

Lab File: VX048525.D

Acq: 07 Nov 2025 16:50




Tgt Ion:225 Resp: 58865


Ion Ratio Lower Upper

225 100

223 61.9 30.6 92.0


227 63.6 32.5 97.4





#95  
Naphthalene  
Concen: 52.576 ug/l  
RT: 13.756 min Scan# 2078  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX110725



Tgt Ion:128 Resp: 506110  
Ion Ratio Lower Upper  
128 100  
127 12.7 10.1 15.1  
129 10.9 8.7 13.1

Manual Integrations  
APPROVED

Reviewed By :John Carlone 11/10/2025  
Supervised By :Semsettin Yesilyurt 11/10/2025



#96  
1,2,3-Trichlorobenzene  
Concen: 50.471 ug/l  
RT: 13.939 min Scan# 2108  
Delta R.T. -0.000 min  
Lab File: VX048525.D  
Acq: 07 Nov 2025 16:50

Tgt Ion:180 Resp: 140560  
Ion Ratio Lower Upper  
180 100  
182 95.1 47.3 142.0  
145 33.2 16.7 50.0

