

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110818\
 Data File : VX005843.D
 Acq On : 08 Nov 2018 13:43
 Operator : JC/MD
 Sample : J5860-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP2-EB-01-1118

Quant Time: Nov 21 00:10:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	112563	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	147408	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	129525	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	63785	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	66316	51.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.50%	
35) Dibromofluoromethane	5.50	113	55031	54.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.78%	
50) Toluene-d8	8.71	98	181437	54.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.86%	
62) 4-Bromofluorobenzene	11.14	95	58450	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
Target Compounds						
16) Acetone	2.45	43	2037	2.924	ug/l	96
95) Naphthalene	13.83	128	735	1.334	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX110818\
Data File : VX005843.D
Acq On : 08 Nov 2018 13:43
Operator : JC/MD
Sample : J5860-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP2-EB-01-1118

Quant Time: Nov 21 00:10:15 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

